AperTO - Archivio Istituzionale Open Access dell'Università di Torino

Application of mass spectrometry to hair analysis for forensic toxicological investigations

This is the author's manuscript

Original Citation:

Availability:

This version is available <http://hdl.handle.net/2318/117256> since 2015-12-29T11:15:33Z

Published version:

DOI:10.1002/mas.21364

Terms of use:

Open Access

Anyone can freely access the full text of works made available as "Open Access". Works made available under a Creative Commons license can be used according to the terms and conditions of said license. Use of all other works requires consent of the right holder (author or publisher) if not exempted from copyright protection by the applicable law.

(Article begins on next page)

APPLICATION OF MASS SPECTROMETRY TO HAIR ANALYSIS FOR FORENSIC TOXICOLOGICAL INVESTIGATIONS

**Marco Vincenti,^{1,2*} Alberto Salomone,¹ Enrico Gerace,¹ and
Valentina Pirro^{1,2}**

¹*Centro Regionale Antidoping e di Tossicologia "A. Bertinaria", Orbassano (TO), Italy*

²*Dipartimento di Chimica, Università di Torino, Torino, Italy*

Running head title: "Hair Analysis for Forensic Toxicology"

Keywords: Hair, Toxicology, Drugs, Mass Spectrometry, Forensic

ABSTRACT

The increasing role of hair analysis in forensic toxicological investigations principally owes to recent improvements of mass spectrometric instrumentation. Research achievements during the last six years in this distinctive application area of analytical toxicology are reviewed. The earlier state of the art of hair analysis was comprehensively covered by a dedicated book (Kintz, 2007a), that represents key reference of the present overview. Whereas the traditional organization of analytical methods in forensic toxicology divided target substances into quite homogeneous groups of drugs, with similar structures and chemical properties, the current approach often takes advantage of the rapid expansion of multiclass and multiresidue analytical procedures; the latter is made possible by the fast operation and extreme sensitivity of modern mass spectrometers. This change in the strategy of toxicological analysis is reflected in the presentation of the recent literature material, which is mostly based on a fit-for-purpose logic. Thus, general screening of unknown substances is applied in diverse forensic contexts than drugs of abuse testing, and different instrumentation (triple quadrupoles, time-of-flight analyzers, linear and orbital traps) is utilized to optimally cope with the scope. Other key issues of modern toxicology, such as cost reduction and high sample throughput, are discussed with reference to procedural and instrumental alternatives.

**Correspondence to:*

Prof. Marco Vincenti,
Dipartimento di Chimica, Università di Torino,
Via Pietro Giuria 5, 10125 – Torino, Italy
Tel.: +39 0116705264
Fax: +39 0116705242
E-mail: marco.vincenti@unito.it

1 I. INTRODUCTION

2

3 The analysis of biological specimens to detect of various types of drugs is consistently requested
4 within forensic investigations as a necessary step to prove exposure to these substances. Although
5 blood and urine remain the matrices of choice to detect recent exposure, hair analysis is gaining
6 ever-increasing importance, as its potentials, limitations, and new application areas are
7 progressively uncovered. Although, for most drugs, any evidence of exposure is erased from blood
8 and urine a few days (or hours) after their intake because of the natural metabolic and excretion
9 body processes, a minute portion of the same drugs and metabolites is ultimately incorporated
10 through several routes (blood capillaries, sweat, and sebum) into the keratin structure of the hair,
11 from which it is hardly removed. This drug-fixing into the hair structure resists hair growth for
12 several months and leads to a potential chronological trace of drug exposure, with farther periods
13 corresponding to the hair segments more distant from the hair root. The continuous improvement of
14 analytical procedures and instrumental technologies allows one to determine the very small amounts
15 of drugs included into the hair at ever-decreasing concentrations, insomuch as there are currently
16 several scientific reporting of drug detection in hair after a single exposure. Progresses in
17 chromatographic and mass spectrometric techniques have both equally contributed to the
18 impressive results achievable nowadays with modern forensic toxicology in hair analysis.

19 On the negative side of hair analysis, there are several potential sources of bias that should
20 always be taken into account in order to produce reliable conclusions from experimental results.
21 First of all, hair is not a homogeneous matrix; thus, careful and representative hair sampling is a
22 necessary prerequisite of correct analysis. Secondly, the degree of hair incorporation depends on the
23 chemical structure and related properties of the drug, including melanin affinity, lipophilicity and
24 membrane permeability (Kintz, 2012a). Third, adsorption of external substances from the
25 environment is occasionally claimed as a source of false-positive results, especially for
26 professionally-exposed subjects and volatile substances, such as, for example, cocaine (LeBeau &
27 Montgomery, 2009). To prevent this inconvenience, preliminary washing of hair samples is always
28 recommended to eliminate any possible interfering substances adsorbed onto the external surface of
29 the hair. Segmental analysis and acceptance of cut-off values often rule out the chance that
30 exogenous drug sources might have contaminated the internal keratin structure. Forth, strong
31 cosmetic hair treatments, such as the use of oxidants or highly basic coloring, might damage the
32 keratin structure to favor release of incorporated substances to finally lead to false-negative results.
33 Also, repeated shampooing might partly wash out the hydrophilic drugs incorporated into the hair.

34 Lastly, the use of cosmetic products such as grease, sprays, and gels might also interfere with hair
35 analysis.

36 A comprehensive review of drug testing in hair was published in a dedicated book edited by
37 Pascal Kintz (2007a). Since then, only one review was published on abused drug analysis in hair
38 samples, in which chapters are organized on the basis of the various classes of substances (Wada et
39 al., 2010). The present review updates Kintz's book, under the perspective of its chapters 10-15
40 (Kintz, 2007a); namely, under application areas point of view. The studies published before 2006
41 are only incidentally cited in the present review, whereas extensive coverage of the papers
42 published in the last six years is presented.

43

44 **II. APPLICATION AREAS OF MASS SPECTROMETRY TO HAIR**

45 **ANALYSIS IN FORENSIC TOXICOLOGY**

46

47 **A. Targeted Drug Screening for Specific Drug Classes**

48 Although screening of psychoactive substances in biological specimens might have a variety of
49 objectives, in terms of target chemical classes and purposes of control, legislation of most countries
50 make a clear distinction between drugs of abuse, whose use is prohibited in almost all
51 circumstances, and pharmaceutical substances, whose use is permitted under medical control and
52 prescription, even though they are occasionally abused. Most routine screening, such as, for
53 example, workplace testing, are addressed to the drugs of abuse only, whereas other psychoactive
54 pharmaceuticals are searched in selected subjects and follow anamnestic indications. In particular,
55 detection of the most-common drugs of abuse in hair samples (Kintz, 2007a; Kintz et al., 2006;
56 Pragst & Balikova, 2006; Wada et al., 2010) is increasingly requested for the retrospective
57 withdrawal control of habitual drug abusers, as well as in other toxicological investigations, such as
58 workplace drug testing [EWDTS (European Workplace Drug Testing Society), 2010; Caplan &
59 Goldberger, 2001], driving re-licensing, occasional or pre-natal exposure to drugs (Kintz, 2007b;
60 Falcon et al., 2012), and post-mortem toxicology. Therefore, although hair samples are routinely
61 collected and analyzed in forensic and toxicological laboratories, most published methods were
62 implemented to determine class-specific groups of compounds, mainly cannabinoids,
63 amphetamines, cocaine, and opiates, even though the analytical procedures have been constantly
64 updated (Kintz, 2007a; Pragst & Balikova, 2006; Wada et al., 2010).

65

1. Classes of drugs

Although the widespread abuse of hashish or marijuana makes the detection of long-term exposure to Δ^9 -tetrahydrocannabinol (THC) by hair analysis extremely important in clinical and forensic contexts, drug inhalation and adsorption from environmental smoke are frequently claimed to justify positive detection of THC and/or its main metabolite. Although a cut-off of 0.1 ng/mg was traditionally recommended, more recently a cut-off of 0.05 ng/mg has been proposed as a suitable level to detect also occasional users (Pragst & Nadulski, 2005) and for confirmation analyses (Cooper et al., 2012).

GC/MS is the technique of choice for THC detection in the majority of published methods, and no particular analytical improvement has been proposed recently, except for the increasing role played by headspace solid-phase microextraction (HS-SPME) procedures. Nadulski and Pragst developed, validated, and routinely applied to driving-ability examination a new method for cannabidiol (CBD) and cannabinol (CBN) with improved sensitivity (Nadulski & Pragst, 2007). The authors used alkaline digestion of 15-30 mg hair aliquots, subsequent liquid-liquid extraction (LLE), automated HS-SPME after in-sample silylation, and GC/MS-SIM (selected ion monitoring) analysis. For THC, a limit of detection (LOD) of 0.012 ng/mg was obtained. A similar procedure without derivatization with 10 mg of hair sample was proposed (Rodrigues de Oliveira et al., 2007), but the reported LOD and LOQ (limit of quantification) values of 0.07 ng/mg and 0.12 ng/mg, respectively, are above the recommended cut-off limits and make the method apparently not appropriate for routine analysis. Emídio et al. developed and validated a method for the determination of THC, CBD, and CBN in hair samples, with HS-SPME combined with GC/MS/MS (ion trap) detection (Emídio et al., 2010). From a 10 mg hair sample, an LOD of 0.031 ng/mg and LOQ of 0.062 ng/mg were obtained; the latter was slightly higher than the cut-off value of 0.05 ng/mg. Nonetheless, from 10 hair samples from cannabis users, an average concentration of 0.056 ng/mg was found; i.e., below the reported LOQ.

Unlike THC, the improvement of existing analytical methods for stimulants, including cocaine and amphetamines, in hair samples has been widely investigated recently. Lee et al. used a standard GC/MS procedure, after derivatization with trifluoroacetic anhydride, in order to study the abundance ratio of methamphetamine (MA) and its metabolite amphetamine (AP) in hair and hence aid the positive results interpretation (Lee et al., 2009). High MA concentrations, together with low AP percentage, were related to severe and chronic drug abuse (Lee et al., 2009). In order to facilitate MA and AP analysis in hair, Miyaguchi et al. developed a simple and fast (1 hour) method for sample preparation named MiAMi (micropulverized extraction-aqueous acetylation-

100 microextraction by packed sorbent) followed by GC/MS (Miyaguchi et al., 2009). The amount of
101 hair sample required for qualitative analysis based on full-scan mass spectra was only 5 mg,
102 whereas 1 mg of a hair specimen was sufficient for amphetamine quantitation. The same group
103 (Miyaguchi et al., 2007) previously published an extraction method on micropulverized hair for the
104 HPLC-MS/MS determination of MA and AP, in which only 2 mg of sample were used.

105 An LC-MS/MS method that used electrospray ionization (ESI) and a triple quadrupole
106 instrument was developed and validated by Chèze et al. to determine AP, MA, 3,4-
107 methylenedioxyamphetamine (MDA), 3,4-methylenedioxymethamphetamine (MDMA, “ecstasy”),
108 3,4-methylenedioxyethamphetamine (MDEA), and N-methyl-1-(3,4-methylenedioxyphenyl)-2-
109 butanamine (MBDB) at low concentration levels, in hair, blood, and urine (Chèze et al., 2007).
110 With 20 mg of decontaminated hair, the experimental LODs ranged from 0.3 pg/mg for MBDB to
111 6.3 pg/mg for MDA, where the recommended cut-off limit for amphetamines is 200 pg/mg (Cooper
112 et al., 2012). The same study reported a real forensic case, in which the high sensitivity of the LC-
113 ESI-MS/MS method was exploited to detect MDMA in a hair sample, after a single oral dose
114 administration to an unaware victim (Chèze et al., 2007). Another recent study reported the
115 development of an LC-MS/MS method for the simultaneous determination of a large set of
116 amphetamine-like anorectics and their metabolites in hair samples (Lee et. al., 2012). The procedure
117 was used to analyze the diffusion of anorectics abuse in Korea. The hair incorporation of MA and
118 AP following controlled administration of MA to seven volunteers was recently studied (Polettini et
119 al., 2012). For these amphetamines, significant dependence (i.e.; linear direct correlation) of drug
120 incorporation on hair melanin content was clearly demonstrated.

121 Besides MALDI-imaging techniques, several traditional methods were recently proposed to
122 determine the most common stimulant drug (i.e., cocaine) in hair samples. The suggested cut-off
123 concentration is 500 pg/mg for cocaine and 50 pg/mg for its main metabolite, benzoylecgonine
124 (BZE) (Cooper et al., 2012). A GC/MS analytical method to quantify cocaine and its main
125 metabolite BZE in hair samples was proposed (Barroso et al., 2008). Despite the simple
126 instrumentation utilized, the method proved to be sensitive and specific, and allowed one to detect
127 20 and 15 pg/mg of cocaine and BZE, respectively, from only 20 mg of sample.

128 Improved sensitivities were obtained with LC-MS/MS methods. Moore et al. developed and
129 validated an LC-MS/MS procedure for the analysis of cocaine and its metabolites (BZE,
130 cocaethylene, and norcocaine) in hair with an atmospheric pressure chemical ionization (APCI)
131 source and a triple quadrupole (QqQ) mass analyzer (Moore et al., 2007). For all analytes, the
132 LOQs (50 pg/mg) and LODs (25 pg/mg) made the method suitable for routine forensic analysis.

133 A simple and fully validated procedure for the qualitative and quantitative determination of
134 opiates in hair was presented (Barroso et al., 2010a). This paper describes a GC/MS method for the
135 analysis of codeine, morphine, 6-monoacetylmorphine (6-MAM), 6-acetylcodeine, and tramadol in
136 20 mg hair samples. The presence of the screened analytes was demonstrated in several real cases,
137 among which it was proven that tramadol is occasionally abused by opiate addicts. A combination
138 of opiates, cocaine and metabolites was simultaneously screened with an LC-ESI-MS/MS method
139 (Huang et al., 2009). This method was fully validated and used in the analysis of 79 authentic hair
140 samples, and demonstrated that a multiresidue approach might screen different classes of drugs at
141 the same time.

142 The objective of cost reduction of workplace drug screening with a multiresidue UHPLC-
143 MS/MS strategy was explicitly cited in a recent study (Di Corcia et al., 2012). Thirteen analytes,
144 including opiates, cocaine, amphetamines, THC, buprenorphine, methadone, and a few metabolites,
145 were simultaneously screened in hair samples. A simple sample preparation combined with
146 multiclass analysis and fast chromatographic separation allowed one to obtain high sample
147 throughput, together with excellent sensitivity and selectivity; the procedure is valuable for large
148 sample workload and reduced costs of analysis (Di Corcia et al., 2012).

149

150 **2. New instrumental set-up**

151

152 A new and original study, aimed to the direct detection of MA in hair samples, used imaging
153 mass spectrometry (IMS) to perform micro-segmental analysis (Miki et al., 2011). In practice, a hair
154 shaft was affixed to a carbon tape and manually cut lengthwise to produce micro-incisions at
155 extremely close range with a razor and a microscope. After matrix deposition, IMS was obtained by
156 matrix-assisted laser desorption ionization time-of-flight (MALDI-TOF) and MALDI-Fourier
157 transform ion cyclotron resonance (FTICR) mass spectrometers. IMS produced a barcode-like
158 image of methamphetamine on the longitudinal sections of the hair shafts, obtained from a
159 methamphetamine chronic abuser, to indicate the periods and sequence of single MA intake (Miki
160 et al., 2011). This innovative approach appears to be very promising in the forensic context, as soon
161 as the hair sample preparation can be automated.

162 A MALDI hybrid triple quadrupole linear ion trap (QqQ_{LIT}) instrument, equipped with a high
163 repetition rate laser, was recently proposed to perform MS imaging on a single hair, for cocaine
164 consumption monitoring (Porta et al., 2011). The QqQ_{LIT} technology was exploited to
165 simultaneously achieve target quantification with selected reaction monitoring (SRM) mode of
166 operation and sensitive MS³ acquisition for confirmatory analysis. MS imaging of intact single hair

sample offers much higher resolution than segmental LC-MS/MS analysis, and only depends on the distance between two adjacent laser shots. However, the minimal 0.1 mm resolution allowed by the laser and corresponding to a theoretical average hair growth of less than 12 h, has currently little significance in forensic investigations, because the complex nature of keratin incorporation mechanisms and the heterogeneous physiology of hair growth might provide a much larger bias on chronological assessment. Single washed hair samples were fixed onto a stainless steel MALDI plate with a double-sided adhesive foil, and the MALDI matrix was manually sprayed. First, an imaging profile of cocaine and metabolites was acquired in the SRM mode over the whole hair length, taking into account that highly selective transitions should be chosen to compensate for the absence of chromatographic separation. Secondly, MS/MS and MS³ experiments were performed for confirmatory identification. The entire analytical workflow is represented in Figure 1 (Porta et al., 2011). An experimental LOD of 5 ng/mg allows for cocaine detection in the hair of chronic abusers, but not after a single administration. Moreover, scaling down to single-hair analysis makes the choice of representative sampling crucial, unless a further source of uncertainty and bias is introduced.

In order to detect all cocaine metabolites, including the highly polar ecgonine methyl ester (EME), Quintela et al. developed and validated a hydrophilic interaction liquid chromatography (HILIC) method to be coupled to tandem mass spectrometry (Quintela et al., 2010). With as little as 10 mg of specimen, experimental LODs were better than or equal to 1 pg/mg. The recognition of ultra-trace amounts, at low pg/mg of hair, appears to be crucial for EME detection, because this metabolite has a very low incorporation rate into the hair shaft (Quintela et al., 2010). Recently, Thibert et al. adopted a clean-up process for hair extracts based on molecularly imprinted polymers (MIP) selective for cocaine and BZE, followed by LC-MS/MS analysis (Thibert et al., 2012). An LOD lower than 70 pg/mg was reached for both molecules, slightly above the suggested cut-off value for BZE (Cooper et al., 2012).

A fast, but only qualitative, screening method to detect cocaine and its metabolites from hair samples used MALDI-TOF technology (Vogliardi et al., 2009). The whole hair sample preparation for MALDI analysis turned out to be significantly more rapid and simpler than for GC/MS, and data acquisition in MALDI analysis is, in turn, much quicker than in GC/MS. Thus, the MALDI technique shows clear advantages over the traditional approach, when fast screening of a large number of sample is required, as the same authors subsequently demonstrated by validating and testing the new method on 304 real hair samples (Vogliardi et al., 2010). High throughput and fit-for-purpose principles both find neat application in this well-designed example of fast cocaine screening in hair samples.

201

202 **B. Non-Targeted Drug Screening**

203 Most of the methods cited in the preceding chapter are specific for a class of drugs. The highly-
204 focused objective of these analytical investigations allows one to optimize sample treatment, clean-
205 up, and chromatographic conditions in order to maximize recoveries and to meet the acceptance
206 criteria of sensitivity and specificity that lead to accurate quantification for confirmation. However,
207 their applicability in general drug screening for acute intoxication and autopsic analysis appears
208 limited. In practice, these class-specific procedures find useful application when the target analytes
209 can be predicted in advance; for example, in heroin overdose fatalities.

210 In the daily activity of forensic laboratories, it is quite frequent that the target analytes cannot be
211 foreseen, so that a wide range or "general" screening of abused drugs is commonly required, as it
212 occurs in most acute intoxications and post-mortem investigations. Whenever a multitude of
213 candidate substances might represent the cause of intoxication or death, several analytical
214 procedures are likely to be used on the collected biological specimen, with a direct impact on time,
215 costs, and efficiency. For this reason, comprehensive screening procedures of multiclass drugs are
216 progressively introduced into the analytical practice, even if the method conditions should
217 compromise on absolute performances, in terms of recovery, sensitivity, selectivity, and accuracy,
218 due to the presence of target analytes with widely different physico-chemical properties.

219 Even though hair is not the preferred biological matrix to ascertain acute intoxication, hair
220 analysis is frequently commissioned to complete the circumstantial evidences; i.e., to ascertain
221 whether the victim formerly abused drugs or whether any drug was administered to him/her earlier.
222 To meet the high demand for drug screening in hair samples, toxicology laboratories are forced to
223 update their procedures in order to target an ever-increasing number of drugs, but also to achieve
224 rapid, simple, and sensitive testing with reduced sample preparation and fast instrumental
225 processing so as to increase the overall sample throughput.

226

227 **1. Methods based on GC separation**

228

229 Cordero and Paterson developed and validated a GC/MS protocol to simultaneously quantify
230 fourteen compounds, including some amphetamines, opiates, cocaine and metabolites, diazepam,
231 and metabolite (Cordero & Paterson, 2007). A two-step derivatization with *N*-methyl-bis
232 trifluoroacetamide (MBTFA) and *N*-methyl-*N*-trimethylsilyltrifluoroacetamide (MSTFA) + 1%
233 trimethylchlorosilane (TMCS) was used. Due to the general availability of GC/MS instrumentation
234 in toxicology laboratories and the good analytical performances obtained in this study, this method

might find useful application in a variety of clinical and forensic investigations. An interesting evolution of the preceding method was developed by Aleksa et al., that used an HS-SPME-GC/MS to improve the protocol sensitivity (Aleksa et al., 2012). Seventeen drugs of abuse, including opiates, cocaine, amphetamines, and opioids, were reacted with the same derivatizing agents described above, and were detected at LOD levels of 130-200 pg/mg. The method was validated with only 5-10 mg of hair, and was especially useful for newborn, when sample availability is limited. Among the wide screening protocols to detect common panels of abused drugs, only few comprehensive methods include THC, regardless of the instrumental technique used. Merola et al. developed a new procedure with HS-SPME and GC/MS to provide concomitant determination of several drugs of abuse, including THC, but not opiates (Merola et al., 2010). A different approach and derivatizing agent, namely heptafluorobutyric anhydride (HFBA), was used (Wu et al., 2008a) to develop a GC/MS method for the simultaneous determination of amphetamines, opiates, ketamine, and their metabolites at LODs of 30-80 pg/mg. The same group (Wu et al., 2008b) improved selectivity and sensitivity to reach the low pg/mg level with electron capture negative chemical ionization (NCI) to detect the same derivatives.

TOF mass analyzers have also been used for broad drug screening in hair samples that take advantage of their full-scan and high-resolution capabilities. In an original study, Guthery et al. developed a GC×GC-TOF-MS procedure to provide a comprehensive qualitative drug screening of hair samples (Guthery et al., 2010). The hair extracts were subjected to derivatization with N-methyl-N-(tert-butyl-dimethyl)trifluoroacetamide (MTBSTFA), which efficiently reacted with a broad range of multiclass analytes, including opiates, opioids, cocaine, and benzodiazepines. The 2D chromatographic plot obtained from GC×GC separation offers an extremely clear depiction of the drug content profile, as Figure 2 demonstrates (Guthery et al., 2010).

258

259 **2. Methods based on LC separation**

260

Although the latter GC/MS methods proved suitable to test simultaneously several classes of drugs of abuse, the need to further expand the assortment of target substances within the same protocol, without the necessity of derivatization, made the highly flexible LC-MS/MS approach to multiclass screening more profitable and more widely used than GC/MS. An LC-APCI-MS/MS method was developed, validated, and applied for the retrospective multi-parameter evaluation and distribution of eleven drugs of abuse in hair samples (Klis et al., 2007). Subsequently, ESI was more extensively utilized than APCI, as in a validated LC-MS/MS method (Miller et al., 2008) that was aimed at the simultaneous quantification of amphetamines, opiates, cocaine and metabolites,

269 and diazepam and metabolites (17 compounds) in post-mortem hair samples. The use of an ion trap
270 mass spectrometer was suggested (Bucelli et al., 2009) to screen 16 drugs of abuse in human hair.
271 Despite the flexibility of ion traps to provide an easy switch between full-scan and MS/MS or MSⁿ
272 modes of data acquisition, the method was not expected to provide accurate quantitative
273 determinations, and needs further validation and a more extended panel before it might find
274 widespread application in forensic laboratories. Di Corcia et al. proposed a simple sample
275 extraction and direct injection into a UHPLC-MS/MS system to avoid solid-phase and liquid-liquid
276 extraction and to achieve fast and simultaneous detection of the most common drugs of abuse,
277 including THC, in hair samples (Di Corcia et al., 2012). Lendoiro et al. developed and validated an
278 LC-MS/MS multiclass screening method for the simultaneous detection of 35 substances, including
279 THC, and the other most common drugs of abuse (opiates, amphetamines, cocaine, LSD, ketamine
280 and scopolamine) and pharmaceuticals (benzodiazepines, antidepressants and hypnotics) (Lendoiro
281 et al., 2012). For the entire range of investigated molecules, LODs were 0.2-50 pg/mg and achieved
282 better cut-offs than Society of Hair Testing (SOHT) recommendations. In the screening mode of
283 operation, only one SRM transition per compound was used, but the same procedure could be
284 transformed into a confirmation method with addition of further target-specific SRM transitions
285 (Lendoiro et al., 2012). Confirmatory analysis of ketamine and norketamine was the objective of a
286 recent study that used molecularly imprinted solid-phase microextraction (MISPE) followed by LC-
287 MS/MS (Harun et al., 2010).

288 An effective method to check for the presence of a wide set of drugs in hair samples was
289 proposed (Nielsen et al., 2010), to validate an analytical method for the simultaneous screening and
290 quantification of 52 common drugs and pharmaceuticals in hair with the novel UHPLC-TOF-MS
291 technology. The panel of compounds was very broad because it included common drugs of abuse,
292 benzodiazepines, analgesics, antidepressants and antipsychotics. The authors underlined, however,
293 that the TOF instrument was less useful than a triple quadrupole on low-mass molecules, such as
294 amphetamines, due to the presence of high background noise and the consequent lack of sensitivity
295 and linearity (Nielsen et al., 2010). Similar broad drug screening, based on accurate mass
296 measurements with LC-TOF-MS, was also achieved (Pelander et al., 2008). Also in this study, the
297 potential to conduct a comprehensive screen (35 targeted analytes) was clearly demonstrated;
298 however, the authors recommended more-specific instrumentation for target confirmation analyses.
299 More recently, an automated screening method for the simultaneous identification and quantitation
300 of 30 compounds (including most drugs of abuse, but also a few anabolic steroids, β -agonists and
301 other pharmaceuticals) in hair samples was proposed with fast LC-TOF-MS (Domínguez-Romero
302 et al., 2011). Progressive instrumental upgrading and method refinement provided improved

303 sensitivity with TOF mass spectrometers to allow the detection of the target compounds in the low
304 pg/mg range (LODs ranged from 5 pg/mg to 50 pg/mg).

305 General screening of unknown drugs in hair samples was performed with LC combined with a
306 quadrupole time-of-flight mass spectrometer (LC-QTOF-MS) (Broecker et al., 2012). Systematic
307 collection of CID (collision-induced dissociation) spectra from quadrupole-selected protonated
308 molecule used a data-dependent acquisition mode to create a wide dataset to allow the retrospective
309 investigation of potentially toxic substances present in the hair sample. These CID spectra were
310 compared with an accurate-mass CID spectral library created by the authors that contained more
311 than 2500 toxicologically relevant substances (Broecker et al., 2011; Niessen, 2011). For 24 tested
312 illegal drugs and benzodiazepines, significantly low LODs (3-15 pg/mg) were observed. The
313 method was applied to hair samples from 90 death cases to reveal 212 substances altogether, even if
314 only partial agreement with the positive identifications from urine and blood analysis was found,
315 with many unexpected drugs detected and many reported drugs not detected in hair (Broecker et al.,
316 2012). A similar approach was also applied to hair samples (Liu et al., 2010) with an in-house
317 spectral library of about 800 substances for the identification of unknown drugs. In this case, the
318 hybrid QqQ_{LIT} mass spectrometer produced low-resolution CID spectra from mass-selected
319 precursor ions, with good sensitivity provided by the ion-accumulation capability of ion trap.

320

321 **3. Other methods**

322

323 Besides GC and LC, capillary electrophoretic methods were also occasionally used to separate
324 drugs extracted from hair samples and to introduce them into the mass spectrometer. Capillary zone
325 electrophoresis was combined with TOF analysis (CZE-TOF-MS) to detect four amphetamines,
326 ephedrine, opiates, cocaine, and metabolites in hair samples. The method proved suitable for
327 qualitative screening of all drugs below the prescribed cut-off limits, but was not fully satisfactory
328 for accurate quantification, possibly due to the lack of deuterated internal standards (Gottardo et al.,
329 2007; Gottardo et al., 2012). Also, another method, cation-selective exhaustive injection and
330 sweeping micellar electrokinetic chromatography, proved compliant with the cut-off values in hair
331 (Lin et al., 2007); however this method and the preceding study represented isolated proposals with
332 no further application in common laboratory practice.

333 Quite uniquely, Jackson et al. proposed to eliminate the chromatographic separation with an
334 ambient ionization technique, defined as low temperature plasma (LTP) ionization, to analyze
335 biological specimens, including urine, saliva, and hair extracts (Jackson et al., 2010). The method
336 was tested on 14 drugs that belonged to different classes (amphetamine, BZE, cannabidiol, caffeine,

cocaine, codeine, diazepam, ephedrine hydrochloride, heroin, ketamine, methadone, methamphetamine, morphine, and THC) with a linear ion trap mass spectrometer in the positive-ion mode. However, LOD values were rather high [ng/mL (i.e., 10 ng/mL) to the mg/mL concentrations] and only one real hair sample was analyzed that showed the presence of cocaine. The main focus of the study was on the potential of LTP ionization and its compatibility with miniaturized and portable instrumentation for on-site screening analysis.

C. Rape Drug Detection and Segmental Analysis

Drug-facilitated crime (DFC) is a general term that involves the administration of a drug to perpetrate a variety of crimes, including rape or other sexual assault, robbery, money extortion, as well as the deliberate maltreatment of elderly persons or children under the influence of psychotropic substances [UNODC (United Nations Office on Drug and Crime), 2011]. In most DFCs, the victims are subjected to violence or nonconsensual acts while they are incapacitated by the effects of a drug, often unknowingly ingested. Psychoactive substances used in DFCs might alter the victim's degree of consciousness, state of awareness, judgment, and memory. Such substances can make the victim vulnerable and unable to fight off the attacker, or are used to sedate the victim in order to facilitate his/her transport by the perpetrator. The drugs used for these criminal purposes include benzodiazepines, hypnotics, sedatives and anesthetics, drugs of abuse (such as cannabis, ecstasy, LSD, or heroin), miscellaneous drugs (e.g., scopolamine), and, most frequently, ethanol (Kintz, 2007b). The ideal substance to be dispensed to perpetrate a crime is the one that is readily available, easy to administer, able to rapidly impair consciousness, and produce anterograde amnesia (LeBeau & Montgomery, 2010).

Among DFCs, a drug-facilitated sexual assault (DFSA) occurs when a person is subjected to unwilling sexual acts while incapacitated or unconscious, due to the effects of ethanol, drugs, and/or other intoxicating substances (generally called "rape drugs"). It is highly unlikely that the specific drug used in a DFSA could be determined only by symptoms because the depressant effects of most of these drugs are similar. This factor makes crucial identification of the administered drug in the victim's biological specimen (Negrusz & Gaensslen, 2003). It is well-known that sexual assaults are significantly under- and late-reported. Evidently, any delay to report a DFSA implies that the victim's biological fluids (i.e., blood and urine) are not timely collected, hence rendering more difficult the toxicologists' task to identify the incapacitating agent (LeBeau & Mozayani, 2001). Whenever the natural metabolic processes have eliminated the drug from biological fluids (typically, urine and blood), hair is the most helpful specimen to be investigated (Negrusz et al., 2000; Villain et al., 2004; Negrusz et al., 2001; Scott, 2009; Parkin & Brailsford, 2009).

371 Toxicology laboratories are frequently asked to analyze hair from victims in order to reveal
372 allegedly administered drugs. Further information is furnished by segmental analysis to generally
373 provide chronological setting. Head hair should be collected at least four weeks after the alleged
374 assault [UNODC (United Nations Office on Drug and Crime), 2011]. Since the average growth rate
375 of human scalp hair is about 1 cm/month, administration of a single dose can be confirmed by drug
376 detection in the corresponding segment, with no presence in the preceding and subsequent
377 segments. The expected concentration in hair is generally in the low pg/mg range for most drugs
378 (Kintz, 2007b; Villain et al., 2005). Therefore, the use of highly sensitive instrumental techniques is
379 mandatory in the investigation of DFCs (Kintz et al., 2005; Chèze et al., 2005).

380

381 **1. Methods based on LC-MS/MS with triple quadrupoles**

382

383 The analytical strategy generally starts with a comprehensive screening for benzodiazepines and
384 other drugs commonly involved in DFCs (Parkin & Brailsford, 2009). In recent years, several
385 methods have been proposed for the simultaneous analysis of a large series of pharmaceuticals and
386 drugs with LC-MS/MS with triple-quadrupole instruments. Hegstad et al. developed a multiclass
387 LC-MS/MS method to screen several drugs of abuse on 20 mg of hair, which included 7-
388 aminonitrazepam, 7-aminoclonazepam, 7-aminoflunitrazepam, oxazepam, diazepam, alprazolam,
389 and two benzodiazepines-like z-drugs (zopiclone and zolpidem) (Hegstad et al., 2008). Similarly,
390 Irving and Dickson described an LC-MS/MS procedure to detect sedatives in hair and nail samples
391 (Irving & Dickson, 2007). Nine benzodiazepines, three metabolites, and zopiclone and its
392 metabolite were selected and detected at low level in 50 mg specimen, with LODs that ranged from
393 0.02 pg/mg for zopiclone to 6.00 pg/mg for temazepam. This procedure was applied to the analysis
394 of a hair sample collected 17 months after the occurrence of a DFSA: zopiclone and its metabolite
395 were both detected in different hair segments at concentrations below 7 pg/mg (Irving and Dickson,
396 2007). Our group developed an LC-MS/MS method to detect 14 pharmaceuticals, including 10
397 benzodiazepines, ketamine, scopolamine, zopiclone, and zolpidem with LODs that ranged from 0.2
398 pg/mg for ketamine and zolpidem to 4.0 pg/mg for flunitrazepam (Salomone et al., 2012a). The
399 method was applied to a real case of segmental analysis of the hair from two women, and showed
400 conclusive discrimination between occasional and therapeutic administration (Salomone et al.,
401 2012a). An interesting study of high-resolution segmental analysis on single hairs was developed to
402 investigate the history of a multiple poisoning case with clozapine (Thieme & Sachs, 2012). From
403 1-2.5 mm segments of a single hair, a highly-specific LC-MS/MS procedure allowed to detect as
404 little as 30 fg (1 pg/mg) of clozapine and its metabolite norclozapine. The collection of hair samples

405 after a 165 days interval, combined with the subsequent comparison of drug concentration profiles,
406 provided an accurate estimation of hair growth rate for the selected individuals (Thieme & Sachs,
407 2012).

408 Another LC-MS/MS method was proposed and validated (Xiang et al., 2011) for the
409 quantitative determination of 18 benzodiazepines in 20 mg of hair. LODs ranged from 0.2 pg/mg
410 for estazolam to 5 pg/mg for 7-aminoclonazepam, 7-aminonitrazepam, α -hydroxytriazolam, and α -
411 hydroxymidazolam. This method determined the distribution of estazolam after single dose intake
412 in hair segments of healthy volunteers: estazolam concentrations ranged between 0.56 pg/mg and
413 2.60 pg/mg. Two real cases were studied with the same method, where the victims of DFCs were
414 exposed to clonazepam: the presence of this drug was found in several hair segments from victim
415 #1, and ranged from 1.63 pg/mg to 15.47 pg/mg, and in two segments from victim #2 at 1.31 pg/mg
416 and 11.93 pg/mg concentrations. The metabolite 7-aminoclonazepam was found in hair segments at
417 considerably higher concentrations than the parent drug to demonstrate the opportunity to include
418 also the main metabolites in the benzodiazepine-screening procedure (Xiang et al., 2011). Similarly,
419 Kim et al. established and validated an analytical method to simultaneously detect 27
420 benzodiazepines and metabolites plus zolpidem in hair with LC-MS/MS (Kim et al., 2011). The
421 absolute LODs ranged from 0.005 ng (zolpidem) to 0.5 ng (bromazepam and chlordiazepoxide).
422 The protocol was successfully applied to five forensic cases of either DFCs or benzodiazepines
423 abuse, together with a study on drug incorporation into the hair of rats. As model for human hair, it
424 was proved that rat-hair pigmentation does not have any significant effect on the degree of
425 benzodiazepines (and metabolites) deposition in hair. As a matter of fact, the incorporation rates of
426 other classes of drugs proved to depend on the hair pigmentation, to create ethnic discrimination
427 (Gambelunghe et al., 2007; Poletini et al. 2012). The panel of screened compounds was expanded
428 further by Lendoiro et al., who recently presented a target multi-analyte LC-ESI-MS/MS screening
429 method to detect 35 licit and illicit drugs, including THC, in 50 mg hair samples (Lendoiro et al.,
430 2012). Among the screened substances, zolpidem, zopiclone, and 12 benzodiazepines were detected
431 at LODs that ranged from 2 pg/mg (7-aminoflunitrazepam, flunitrazepam, lorazepam,
432 lormetazepam, nordiazepam, and diazepam) to 10 pg/mg (oxazepam, clonazepam, and tetrazepam).
433 Analysis of seventeen hair specimens from various forensic cases demonstrated the method
434 applicability: several illicit and prescription drugs were tentatively identified, and subsequently
435 confirmed with two SRM transitions in place of one (Lendoiro et al., 2012). In all the cited LC-
436 MS/MS procedures, triple-quadrupole instruments were used, and proved the efficiency of this
437 instrument for wide multi-analyte screening. Similar LC-MS/MS configuration was also used in
438 confirmation methods specifically designed to determine acepromazine (Gaulier et al., 2008),

439 diazepam, nordiazepam and tetrazepam (Laloup et al., 2007), sildenafil (Kintz et al., 2009), and
440 gamma-hydroxybutyrate (Stout et al., 2010) in hair samples, with the purpose to investigate DFSA
441 cases.

442

443 2. High-resolution mass spectrometers

444

445 Mass spectrometers other than triple quadrupoles were used for wide-range screening and
446 confirmation analysis. Mass analyzers with high-resolution capability and accurate mass detection,
447 such as TOF-MS and FTICR-MS, are particularly suited for *a-posteriori* reevaluation of the
448 presence of drugs not considered in a preliminary target analysis. UHPLC-TOF-MS and UHPLC-
449 MS/MS were both used in the development of a specifically validated confirmation protocol, aimed
450 to quantification of triazolam in an emblematic case of DFC (Stybe Johansen and Dahl-Sørensen,
451 2012). In the hair segments (2 cm) that corresponded to the period of the alleged assaults, the
452 presence of triazolam was revealed at a concentration of 1.0 pg/mg. A wider range of analytes was
453 covered in a high-resolution mass spectrometric (HRMS) drug screening with TOF-MS (Pelander et
454 al., 2008). The authors initially developed an analytical strategy for the comprehensive screening of
455 drugs and doping agents in urine, but afterward the applicability of the method to other matrices,
456 such as keratin, was also studied. Although the potential of the LC-TOF-MS approach to obtain a
457 comprehensive drug screening in the hair of drug addicts was clearly proved, the study did not
458 clarify whether the method sensitivity was high enough to detect drug traces from single intake. A
459 UHPLC-TOF-MS screening procedure previously cited (Nielsen et al., 2010) targeted 52 analytes,
460 among which 13 benzodiazepines and metabolites plus zaleplon, zolpidem, and zopiclone. For the
461 screened molecules, homogeneous LODs were obtained that ranged from 10 pg/mg for diazepam
462 and zolpidem, to 40 pg/mg for 7-aminoflunitrazepam and lorazepam. The method was applied to 15
463 autopsy hair samples. In most strands, several benzodiazepines were identified. Although accurate-
464 mass analysis provided comparable selectivity to multiple SRM transitions from a triple-quadrupole
465 instrument, the authors recommended positive findings to be confirmed with an additional method
466 (Nielsen et al., 2010).

467 Orbitrap technology was also used for HRMS multi-target screening of hair samples (Vogliardi
468 et al., 2011). The purpose of this study was to develop and validate a specific method based on LC-
469 HRMS for the simultaneous detection in 50 mg hair strands of 28 benzodiazepines (and
470 metabolites) registered in the Italian market. The "z-drugs" were not included, despite the fact that
471 zolpidem is significantly prescribed among the Italian population. The use of a hybrid linear ion
472 trap-Orbitrap mass spectrometer yielded LODs from 0.5 pg/mg (clotiazepam) to 10 pg/mg

(delorazepam). The validated procedure was applied to hair samples from suspected benzodiazepine consumers and showed its suitability for screening and identification of the target compounds. Indeed, the potential of such an instrument includes full-scan exact-mass measurement of protonated molecule ions and CID fragment ions, isotopic pattern recognition, and ion ratio determination between molecular and fragment ions. Recently, the same group developed and validated an analogous method to detect 28 substances that belonged to several drug classes: cocaine, amphetamines, opioids, benzodiazepines, antidepressant, and hallucinogens. Although a mass spectrometric approach similar to the one previously described was applied, a new and faster extraction process on micropulverized hair allowed reliable identification and quantification of the target drugs on as little as 2.5 mg of hair samples (Favretto et al., 2012). The combined use of full-scan HRMS and CID offered the possibility of retrospective analysis to detect previously untargeted substances.

485

486 **D. Biomarkers for Alcohol Abuse**

487 An important goal of forensic toxicology and clinical medicine is to identify appropriate
488 biological markers of ethanol consumption to evaluate harmful drinking and to ascertain alcohol
489 abstinence (Pragst & Balikova, 2006). The determination of direct ethanol metabolites, such as
490 ethyl glucuronide (EtG), ethyl sulfate (EtS), fatty acid ethyl esters (FAEEs), and
491 phosphatidylethanol (PEth), in different biological matrices currently represents the most-valid
492 strategy to provide unbiased evidence of chronic alcohol abuse. Among direct biomarkers, FAEEs
493 and EtG can be detected in a variety of matrices, including hair (Pragst & Balikova, 2006), and
494 allow one to gain information on alcohol consumption for several months. Hair analysis proved
495 very useful to monitor ethanol use/abstinence, workplace testing, driving license reissue/renewal,
496 child custody, divorce proceeding, withdrawal treatment (Politi et al., 2006), and post-mortem or
497 pre-natal alcohol exposure investigation (Bendroth et al., 2008; Pragst & Yegles, 2008), even if the
498 straightforward interpretation of analytical results has been fiercely questioned and warning about
499 possible intra- and inter-individual variability has been raised (Tagliaro et al., 2011).

500 According to the Society of Hair Testing, analysis of the proximal 3 cm hair segment is
501 recommended in order to avoid partial loss of the analytes. The suggested cut-off values that
502 support a diagnosis of chronic excessive alcohol consumption are 30 pg EtG/mg scalp hair, and 0.5
503 ng FAEEs/mg scalp hair, respectively; both measured in the 0-3 cm proximal segment (Kintz,
504 2012b). Figure 3 shows the effectiveness of the 30 pg EtG/mg cut-off value (horizontal line) to
505 discriminate harmful drinkers (red dots) from low-risk drinkers and non-drinkers (yellow and green
506 dots), whereas the cut-off value (vertical line) commonly adopted for an esteemed indirect

507 biomarker, such as carbohydrate-deficient transferrin (CDT), proved to have reasonable specificity
508 but largely insufficient sensitivity, that led to about 50% false negative results (Pirro et al., 2011a).

509 In contrast to other forensic applications, a very small number of direct alcohol biomarkers
510 represent for toxicologists the target analytes in hair samples, essentially FAEEs and EtG.
511 Consequently, considerable effort has been devoted to their extraction procedures and instrumental
512 arrangements, in order to obtain continuously improving analytical performances and, above all,
513 minimal LOD and LOQ values. In particular, the refinement of existing methods was addressed to
514 (i) increase the laboratory throughput, with a reduction of sample preparation steps, analysis-time,
515 and costs, (ii) increase the sensitivity and detection capability, so as to accurately quantify minute
516 biomarkers hair concentrations (in the pg/mg range), especially to ascertain alcohol abstinence, (iii)
517 improve the method validation protocols in order to consider all potential sources of bias and
518 uncertainty and assure highly reliable analytical results. Under such circumstances, quite traditional
519 instrumentations, such as triple-quadrupole mass spectrometers in the SRM mode, combined with
520 either GC or HPLC, were almost uniquely utilized, because they assure sensitive and accurate
521 quantification of few pre-determined target analytes.

522

523 **1. FAEE determination in hair samples**

524

525 Due to the relatively high volatility and hydrophobicity of FAEEs, HS-SPME-GC/MS is the
526 suggested method-of-choice for the routine determination of ethyl myristate (E14), ethyl palmitate
527 (E16), ethyl stearate (E18), and ethyl oleate (E18:1) in hair extracts (Kintz, 2012b). Among over 20
528 ethyl esters detected in hair, these four selected analytes were found to best reflect harmful drinking
529 (Pragst et al., 2001). In brief, deuterated internal standards are added to decontaminated hair
530 samples, and subsequently extracted overnight with a heptane-dimethylsulfoxide mixture. The
531 hydrocarbon layer is evaporated and the residue dissolved into a phosphate buffer, on which HS-
532 SPME is executed. The SPME fiber is analyzed with GC/MS in the electron impact (EI) ionization
533 mode under SIM conditions.

534 Over time, laboratories interested in testing for FAEEs adopted the protocol (Pragst et al.,
535 2001), with small changes for in-house revalidation of the method and to optimize the experimental
536 conditions, to determine E14, E16, E18 and E18:1, with the purpose to better clarify the reliability,
537 limitations, and potential of FAEE biomarkers in hair to ascertain harmful drinking. Later, Pragst
538 and Yegles presented an updated procedure, with detection limits ranging between 0.003-0.01
539 ng/mg, as determined on 30 mg of hair (Pragst & Yegles, 2008).

540 Kulaga et al. investigated the potential effect of hair pigmentation on FAEEs incorporation into
541 the keratin matrix, and used, for the first time, hair collected from rats as a model matrix for FAEE
542 production and hair inclusion (Kulaga et al., 2009). As expected for neutral and lipophilic
543 molecules such as FAEEs, they concluded that FAEE concentration in hair does not depend on hair
544 pigment. However, they noted that FAEE hair concentration in rats was much lower than that
545 typically observed in humans, at comparable dosage, but still measurable with the analytical
546 protocol adopted. Gareri et al. used the same analytical method to study the potential bias from use
547 of ethanol-containing hair-care products on FAEE concentration measured on hair of a non-
548 compliant female population (Gareri et al., 2011). From the comparison of FAEE and EtG
549 concentrations in the same hair samples, they concluded that use of hair-care products that contain
550 ethanol (even at low percentage; i.e., less than 10% by volume) might affect FAEE levels in real
551 hair samples, with experimental values from approximately 0.5 to 5.0 ng/mg, to possibly produce
552 false non-compliant outcomes (Gareri et al., 2011). Süssé et al. considered the experimental results
553 from as many as 1872 hair samples from forensic cases (Süssé et al., 2012). They deduced that (i)
554 the use of hair spray might affect (increase) detected FAEEs, but not EtG, hair concentrations; (ii)
555 conversely, bleaching and dying hair treatments might affect (decrease) EtG, but not FAEEs, hair
556 concentrations, whereas (iii) hair gel, hair wax, oil, and grease have apparently no influence on
557 FAEEs and EtG.

558 Süssé et al. reported a method with detection limits ranging from 0.008 to 0.026 ng/mg, as
559 measured on 20-50 mg hair samples (Süssé et al., 2010). These analytical features allowed the same
560 authors to accurately measure the basic levels of hair FAEEs in strict teetotalers: the results showed
561 that 20% of abstainers (N=242) presented total a FAEE concentration that exceeded the cut-off
562 values of 0.5 or 1.0 ng/mg, as a function of the examined hair length (Süssé et al., 2012). In
563 contrast, Albermann et al. reported total a FAEEs concentration lower than 0.2 ng/mg for 74% of
564 drivers who self-claim as abstinent (N=160), whereas only 9% of these samples showed a FAEE
565 concentration that exceeded 0.5 ng/mg, among which some true-positive cases (Albermann et al.,
566 2011).

567 Lately, Hastedt et al. updated and revalidated the HS-SPME-GC/MS method originally
568 developed by Pragst et al. (Hastedt et al., 2012; Pragst et al., 2001). Better detection limits were
569 achieved (0.004-0.014 ng/mg), although an expanded time to complete the chromatography cycle
570 was needed (over 30 min) than in the original method.

571 On the other hand, Caprara et al. and Kulaga et al. described a different analytical strategy to
572 quantify a larger number of FAEEs in neonatal human and animal hair samples, with ethyl
573 heptadecanoate as the internal standard (Caprara et al., 2006; Kulaga et al., 2006). They adopted

solid-phase extraction (SPE) in place of SPME to isolate FAEEs from hair extracts, followed with either GC/MS or GC/MS/MS in the chemical ionization (CI) mode, with isobutene as the ionizing gas. Unlike EI, which typically leads to an extensive, but nearly identical, fragmentation patterns for all FAEEs, CI virtually generates only the protonated molecule ion, which is distinctive for each ethyl ester. MS/MS performed in an ion-trap device clearly adds specificity to the method. In choosing CI rather than EI for the analysis, good detection limits were achieved (<2.4-6.6 pg/mg), only recently outperformed, as reported above. FAEE-analysis profile was also expanded to six, rather than four esters, by adding ethyl laureate (E12), and ethyl palmitoleate (E16:1), and cumulative analysis of all commonly occurring FAEEs in biological matrices was suggested to provide a more effective way to identify neonates with suspected pre-natal exposure to ethanol (Caprara et al., 2006; Kulaga et al., 2006).

By taking advantage of the preceding experiences, Zimmermann and Jackson proposed a globally revisited analytical method to determine five FAEEs (E12, E14, E16, E18, and E18:1) in human hair, and aimed to combine high sample throughput with analytical performances (Zimmermann & Jackson, 2010). Instead of extensive washing procedures and an overnight extraction, followed by a 20 or 40 min-long GC/MS analysis, they proposed a method that required a total analysis-time of less than 1 hour, including a 15-minute sonication-extraction procedure, followed with HS-SPME-GC/MS/MS, performed with a quadrupole ion trap. Positive-CI combined with MS/MS provided high selectivity and detection capability, despite a total chromatography cycle time of only 9 min. The limits of detection ranged from 0.002 to 0.030 ng/mg (Zimmermann & Jackson, 2010). More recently, Politi et al. developed a sensitive (LOQ = 0.01 ng/mg for each FAEE) analytical protocol for FAEE determination in hair, based on the widespread SPE and GC-EI-MS techniques (Politi et al., 2011). These techniques are affordable by any toxicology laboratory - a key feature to make this determination routine. Four ethyl esters (E14, E16, E18, and E18:1) were chromatographically separated in about 16 min. Faster chromatographic analysis is prevented by the need to obtain well-separated peaks and no interferences, because FAEEs share most EI fragment ion signals.

601

602 **2. EtG determination in hair samples**

603

EtG is a polar substance with a relatively low molecular weight. Due to these attributes, a single-stage of MS analysis is generally insufficient to comply with the low detection limits in the pg/mg range required for its determination in a complex keratinic matrix. Either GC/MS/MS after derivatization or LC-MS/MS are the methods of choice for EtG determination in hair samples; both

608 are capable to minimize the background interferences and to achieve the optimal detection
609 capability and high specificity, required in toxicological and forensic applications. The limitations
610 of the simple GC-EI-MS approach for EtG determination in hair samples is indirectly demonstrated
611 by a recently developed method, which barely reaches an LOD of 100 pg/mg with as much as 100
612 mg of hair in the SIM acquisition mode (Álvarez et al., 2009). This LOD value is considerably
613 higher than the established cut-off limit to evidence alcohol use disorders (Kintz, 2012b).

614 Although the simplicity and progressively increasing performances of ESI make the LC-MS/MS
615 approach prevalent nowadays, several highly-efficient GC/MS/MS methods have been recently
616 developed and updated to fulfill the legal requisites entailed by regulations of many countries. Paul
617 et al. proposed and validated a GC/MS/MS procedure, with EI, triple-quadrupole mass
618 spectrometer, and SRM (Paul et al., 2008). The instrumental analysis followed water extraction,
619 SPE purification, and derivatization with N,O-bis[trimethylsilyl]trifluoroacetamide (BSTFA). An
620 LOD of 5 pg/mg was obtained, and the chromatography cycle was 7.30 min (Paul et al., 2008). A
621 modified protocol was reported by the same group in 2011, in which the SRM transitions for GHB
622 were added to the same analytical protocol, so as to detect EtG and GHB within a single
623 chromatographic procedure. This combination is of importance in the investigation of drug-
624 facilitated crimes, because GHB and alcohol are frequently combined substances used to perpetrate
625 these offences (Paul et al., 2011).

626 Large-volume injection (LVI) GC-EI-MS/MS was proposed to increase the sensitivity in hair
627 EtG determination. The authors suggested that, upon optimization of streaming volume, purge time,
628 programmable temperature vaporization (PTV) temperature, injection volume, and other
629 experimental parameters, the detection capability can be improved by at least 1–2 orders of
630 magnitude, as compared to the traditional approach (Shi et al., 2010). Even so, they estimated an
631 LOD value of 5 pg/mg, with 20 mg of hair, as in the method previously cited (Paul et al., 2008).

632 NCI with methane was proposed as an alternative to EI within a GC/MS/MS protocol in order
633 to improve analytical sensitivity (Kharbouche et al., 2009). A high cross-section for electron capture
634 was induced by double derivatization of EtG with perfluoropentanoic anhydride. The choice of two
635 selective SRM transitions allowed high specificity in the analyte identification, despite the fast GC
636 protocol. The estimated LOD value for EtG was 3 pg/mg, with 30 mg hair. This analytical protocol
637 was adopted to analyze rat hair samples, so as to evaluate any potential hair-pigmentation effect on
638 EtG incorporation into the keratin matrix. No statistically significant differences were observed
639 between pigmented and non-pigmented hair (Kharbouche et al., 2010). The same group also
640 evaluated the possible effect of hair-care products on EtG incorporation (Sporkert et al., 2012).

641 A similar GC-NCI-MS/MS approach was also proposed (Agius et al., 2010), but the final EtG
642 enrichment was performed with HS-SPME, after derivatization with heptafluorobutyric anhydride.
643 HS-SPME injection yielded further clean-up in the headspace extraction and accumulation of the
644 derivatized analyte on the fiber, with a significant decrease of the LOD value: 0.6 pg/mg, with 10
645 mg hair (Agius et al., 2010). No further improvements were recorded with nonafluorobutyrric
646 anhydride for EtG derivatization instead of heptafluorobutyric anhydride. A similar LOD value (0.7
647 pg/mg) was also achieved (Kerekes et al., 2009), from 20-30 mg of hair, although NCI was
648 followed by single MS detection, with two monitored ions in the SIM acquisition mode. The EtG
649 concentration was in hair collected from several parts of the body for the same subjects determined
650 and compared (Kerekes et al., 2009).

651 Even though the GC/MS/MS procedures for EtG determination in hair proved to be accurate and
652 highly sensitive, especially when electron capture was used, most toxicology laboratories currently
653 prefer LC-MS/MS, because the sample-purification steps are easier and no derivatization is needed.
654 These simplifications substantially contribute to reduce the work and time for sample processing - a
655 crucial aspect in laboratories with high sample load.

656 Although the first LC-MS/MS method dates back to 2002, the real bloom of analytical
657 procedures based on LC separation and ESI started in 2007. ESI is commonly used in the negative
658 ion mode because of the acidic nature of EtG. Among the large variety of proposed methods,
659 different sample-preparation steps, amounts of analyzed hair, and chromatographic conditions were
660 adopted that yielded excellent selectivity and LOD values between 1 and 10 pg/mg. The proposed
661 methods proved highly reliable to support evaluation of harmful drinking (Morini et al., 2009;
662 Marques et al., 2010; Pirro et al., 2011a; Agius et al., 2012); monitor total abstinence (Kronstrand et
663 al., 2012); study hair-treatment effects that possibly lead to partial disappearance of EtG from the
664 keratin matrix (Morini et al., 2010); and evaluate the effect of individual and external parameters on
665 EtG accumulation and persistence in hair, so as to define the boundaries and to exclude conditions
666 for its use as a biomarker of ethanol consumption (Pragst et al., 2010, Süsse et al., 2012; Hastedt et
667 al., 2012; Cabarcos et al., 2012; Pirro et al., 2011b). Almost all these methods made use of triple-
668 quadrupole instruments to monitor SRM transitions between the deprotonated molecule ion of EtG
669 (m/z 221) and two fragment ions (m/z 75, 85). For EtG determination in hair samples, a detailed list
670 of MS conditions used in the cited articles is presented in Table 1. Conditions for GC and LC
671 methods are reported.

672 The major critical point of hair EtG determination with LC-ESI-MS/MS methods is the possible
673 occurrence of matrix effects (Matuszewski et al., 2003; Peters & Remane, 2012), especially when a
674 non-polar stationary phase is used for LC separation; i.e., EtG eluted quickly. In any case, matrix

675 effects need to be carefully evaluated during method development and validation (Lamoureux et al.,
676 2009; Tarcumnicu et al., 2010; Cabarcos et al., 2012). Several authors proposed to modify
677 previously developed analytical protocols in order to improve sample clean-up (essentially with
678 SPE) and consequently reduce ion-suppression and increase the signal-to-noise ratio. Politi et al.
679 reported an extreme case, in which SPE clean-up proved useful to decrease ion suppression and
680 identify EtG traces in three hair segments of a known alcoholic, 27 years after his death (Politi et
681 al., 2008).

682 Carbon-blend SPE columns, based on porous graphitic material, are suitable for the extraction
683 of EtG from complex matrices such as hair (Lamoureux et al., 2009). Besides, the use of silica
684 HPLC columns, in place of reversed-phase columns, allows one to elute EtG with a mobile phase
685 that contain a high percentage of organic solvent and low percentage of water, to improve ESI
686 efficiency (Lamoureux et al., 2009). An alternative strategy, proposed previously, made use of post-
687 column addition of acetonitrile to the eluate. For the same goal, the use of HILIC coupled to mass
688 spectrometry was proposed and successfully applied for the analysis of EtG in hair (Kintz et al.,
689 2008; Tarcumnicu et al., 2010). Recently, Albermann et al. proposed a modification of their method
690 for EtG determination in hair, aimed to solve occasional chromatographic problems during routine
691 work (Albermann et al., 2012). The new method used an LC column with stationary phase
692 composed of 100% porous graphitic carbon, instead of the reversed-phase LC column previously
693 used (Albermann et al., 2010). However, the new method led to a longer analysis-time and slightly
694 higher LOD value (1.7 pg/mg) than previously reported (1 pg/mg).

695

696 **E. Antidoping Analysis**

697 In amateur and professional sport, the artificial enhancement of the athletic performance with drugs
698 or forbidden practices is subjected to systematic control. Doping controls in sport events are
699 managed by the World Anti-Doping Agency (WADA), through tests of urine or blood sample
700 collected from athletes in and out of competition. Hair analysis is currently not accepted as an
701 alternative to urine and blood analysis for doping control (WADA, 2012) for several reasons,
702 including its heterogeneous nature, the uncertainty on mechanisms that regulate drug incorporation,
703 and alleged diversity in occurrence and extent of these mechanisms among different ethnic groups
704 (Kintz, 2007a). However, it is evident that hair analysis would be of practical significance in several
705 circumstances, because it provides complementary information with respect to urine and blood
706 testing (Gaillard et al., 2000; Kintz, 2003). In fact, urine and blood analyses detect recent drug
707 intake, but cannot distinguish between chronic use or single exposure, whereas hair analysis can
708 offer this distinction (Kintz, 2007a). Indeed, doping agents are partially excreted in the sweat and

709 incorporated into growing hair, and remain stable for long periods of time. Thus, hair testing might
710 offer a large detection window to allow long-term detection of prohibited drugs and a retrospective
711 estimation of the intake period. This opportunity appears to be particularly important to detect out-
712 of-competition doping treatments; for example, in the athletes who use anabolic substances during
713 the training period and interrupt their intake long before the start of competition (Kintz et al., 2000).

714 The need for equal conditions for all athletes and the freedom for them to make use of whatever
715 cosmetic treatment, including complete body shaving, prevented any official use of hair analysis in
716 doping controls. For example, the incorporation rates of prohibited drugs was proved to depend on
717 the hair pigmentation, to possibly lead to unequal conditions in doping control, and consequently
718 ethnic discrimination (Gambelunghe et al., 2007). In addition, hair washing, discoloring, and tinting
719 appear to influence the drug concentration measured in hair, as well as their distribution within the
720 hair longitudinal axis (Kintz, 2003). Nevertheless, the potential of hair analysis for antidoping
721 purposes is so strong that several analytical methods were developed in the past to detect doping
722 agents in hair specimens (Kintz, 2007a).

723 In the last five years, a limited number of analytical methods was additionally developed, with
724 the main objective to measure anabolic androgenic steroid (AASs) level in human hair with gas or
725 liquid chromatography coupled to high-resolution mass spectrometry or tandem mass spectrometry
726 (Anielski, 2008; Gambelunghe et al., 2007; Deshmukh et al., 2010). A complex challenge was
727 addressed to exploit hair analysis to distinguish between natural production of endogenous steroids
728 (e.g., testosterone) and exogenous uptake of the same steroids (Shen et al, 2009; Pozo et al, 2009,
729 Deshmukh et al. 2012). AASs represent the most frequently detected class of substances in out-of-
730 competition doping testing, because they are widely abused to increase strength and lean body
731 mass, and also lead to reduced recovery periods (Thieme et al., 2000; Pozo et al., 2009; Deshmukh
732 et al. 2010). AASs are also utilized to speed up muscle growth in cattle and horses (Gaillard et al.,
733 1999; Anielski, 2008). AAS detection in hair specimens is more widely accepted in animals than in
734 humans; it still receives a considerable interest in the scientific research, as demonstrated recently
735 (Nielen et al., 2011) in which hormone and veterinary drug screening and, in general, forensic
736 investigations can benefit from the recent developments in desorption electrospray ionization
737 (DESI) mass spectrometry (Takáts et al., 2004). DESI allowed bovine hair analysis with very
738 limited sample preparation: after rapid ultrasonic liquid extraction, a few microliters of supernatant
739 were deposited onto a glass or PTFE surface, and dried, before DESI analysis. Full-scan and MS³
740 experiments with a linear ion trap MS were used to detect estradiol benzoate, testosterone
741 cypionate, and testosterone decanoate to demonstrate the general feasibility of rapid screening and

742 detection of anabolic steroid intact esters. Quite high levels (300–800 g/kg) of steroid esters are
743 generally present in hair samples from controlled and illegally treated animals (Nielen et al., 2011).

744 Anielski published a method to detect anabolic steroids and their esters in hair material
745 (Anielski, 2008). After hair extraction (sonication, methanol, 4h, 50°C) and LLE raw purification, a
746 refined sample clean-up procedure was carried out with HPLC. The residue was subsequently
747 derivatized, and the trialkylsilyl steroid derivatives were analyzed with GC/HRMS and GC/MS/MS
748 to yield LOD values between 0.1 and 5.0 pg/mg. The method was successfully applied to real
749 samples to detect testosterone propionate in the hair of treated horses. Unlike a urine specimen, the
750 parent drugs instead of their metabolites are more frequently detected in hair samples.

751 The goal to determine the natural occurrence of anabolic steroids in hair, and their biological
752 variability within selected populations of individuals, led Shen et al. to develop a specific
753 GC/MS/MS method to simultaneously identify testosterone, epitestosterone, androsterone,
754 etiocholanolone, and dehydroepiandrosterone (DHEA) (Shen et al., 2009). After alkaline digestion
755 (NaOH, 1N, 10 min, 90°C), LLE extraction with diethylether was carried out. The extracts were
756 derivatized with a MSTFA/iodotrimethylsilane/dithioerythritol mixture (10:5:5, v/v/w) and
757 analyzed with GC/MS/MS/SRM. The method was fully validated and applied to 80 subjects (39
758 males, 30 females, 11 children). The sensitivity of the method (LODs between 0.1 and 0.2 ng/mg)
759 proved sufficient to determine the physiological concentration ranges for the selected endogenous
760 steroids in the three populations, so as to address the interpretation of non-compliant results from
761 alleged steroid abuse (Shen et al., 2009).

762 An analogous GC/MS/MS method was previously developed for the simultaneous
763 determination of methyltestosterone, nandrolone, boldenone, fluoxymesterone, cocaine, and
764 benzoylecgonine, and was applied to hair samples from seven athletes with reported AAS abuse
765 (Gambelunghe et al., 2007). All anabolic steroids included in the procedure could be detected with a
766 10 pg/mg detection limit. Methyltestosterone, nandrolone, boldenone, and fluoxymesterone were
767 detected in real samples between 12 pg/mg and 37 pg/mg.

768 A specific LC-MS/MS method was developed for the clinical determination of testosterone
769 undecanoate in hair samples (Pozo et al., 2009). After decontamination and incubation in a tris(2-
770 carboxyethyl)phosphine hydrochloride solution (1.5h, 50°C) within an ultrasonic bath, a double
771 LLE with n-pentane was performed. The reconstituted residue was injected into an LC-MS/MS
772 triple-quadrupole mass spectrometer with APCI source. The proposed fragmentation pattern of
773 testosterone undecanoate is shown in Figure 4. The method was applied to hair samples from three
774 patients treated with testosterone undecanoate and led to detected concentrations of 0.4, 1.6 and 8.4
775 pg/mg.

776 The group of Naughton and coworkers recently published two sensitive LC-MS/MS methods
777 for the detection of various steroids in human hair samples. In the first study, nandrolone and
778 stanozolol were determined in human hair for the first time (Deshmukh et al., 2010). After alkaline
779 digestion (NaOH, 1M, 15 min, 95°C) and LLE with n-pentane, the extracts were injected into a
780 UHPLC system interfaced to a triple-quadrupole mass spectrometer with ESI source. From 20 mg
781 aliquots of hair, LOD values of 0.5 pg/mg for stanozolol and 3.0 pg/mg for nandrolone were
782 obtained. The method was successfully applied to 19 hair samples previously tested positive with
783 ELISA (enzyme-linked immunosorbent assay) screening to reveal seven false-positive results and
784 to confirm one sample positive to nandrolone (14.0 pg/mg) and eleven samples positive to
785 stanozolol (concentration range 5.0÷86.3 pg/mg). The second study used a similar experimental
786 approach to develop a highly sensitive LC-MS/MS method to determine testosterone and
787 epitestosterone in hair samples (Deshmukh et al., 2012). The natural abundance ratio of these
788 steroids (T/E) is significantly modified by any exogenous intake of testosterone. Low LOD values
789 (0.1 pg/mg and 0.25 pg/mg for testosterone and epitestosterone, respectively, from 50 mg of hair)
790 allowed the determination of physiological testosterone and epitestosterone levels on 75 real
791 samples, which ranged from 0.70 to 11.81 pg/mg in males and 0.33 to 6.05 pg/mg in females
792 (testosterone); from 0.63 to 8.27 pg/mg in males and 0.52 to 3.88 pg/mg in females
793 (epitestosterone).

794

795 **III. FUTURE TRENDS OF HAIR ANALYSIS**

796

797 **A. New Designer Drugs**

798 In the last years, forensic laboratories have been challenged worldwide with the puzzle to detect
799 a variety of new psychoactive substances. These compounds, occasionally called either designer
800 drugs or legal highs, have little or no previous history of medicinal use. Amphetamine derivatives,
801 with backbone molecular structures of piperazine and cathinone, initially represented the main
802 group of designer drugs. Almost at the same time, another wide group of compounds, known as
803 synthetic cannabinoids, appeared on the black market. A very rapid proliferation of new
804 psychoactive substances of these and other classes followed in recent years. Although most of the
805 latest drugs act as central nervous system (CNS) stimulants, their assorted chemical structures range
806 from derivatives of pipradrol, ketamine, phencyclidine, arecoline, aminopropylbenzofuran, ring-

807 substituted aminoindans, thiophenyl bioisosteres of methamphetamine, as well as compounds
808 structurally related to cocaine (King & Kicman, 2011).

809 Reference standards for most designer-drugs metabolites are not yet commercially available.
810 That lack of standards poses a serious and longstanding challenge to toxicological laboratories to
811 develop analytical procedures to detect their presence in a variety of biological matrices.
812 Nevertheless, several analytical methods were developed to determine some of these new
813 psychoactive substances and/or their metabolites in either blood or urine (Moran et al., 2011;
814 Grigoryev et al., 2011; Dresen et al., 2011; Teske et al., 2010; Wintermeyer et al., 2010; Beuck et
815 al., 2011; Kraemer et al., 2009; Sobolevsky, Prasolov, & Rodchenkov, 2010). On the other hand,
816 only very few studies have been published to date that describe protocols for the detection of these
817 substances in hair samples.

818 In a Letter to the Editor, Torrance and Cooper reported the detection of mephedrone in hair
819 samples at 4.2 and 4.7 ng/mg concentration with a ISO/17025 accredited method, but details on the
820 analytical method used and comments or interpretations of results were not included (Torrance &
821 Cooper, 2010). A specific and accurate method for mephedrone detection with GC/MS was
822 proposed (Martin et al., 2012). The study reported that 67 real hair samples were analyzed - 10 were
823 found positive, with mephedrone concentrations from 0.2 to 313.2 ng/mg; 8 of them were below 6
824 ng/mg. As for amphetamines, it is deduced that mephedrone concentrations in hair are likely in the
825 ng/mg, not pg/mg, at least in the cases of repeated abuse (Martin et al., 2012). Detection of
826 piperazine-like compounds was meticulously described (Barroso et al., 2010b). This study was the
827 first to report the development and validation of an analytical method for the determination of three
828 phenylpiperazines in hair samples. Trimethylsilyl derivatives were determined with GC/MS, and
829 the method was applied to autoptic samples and samples collected from subjects under psychiatric
830 evaluation.

831 The first multiclass screening for synthetic cannabinoids in hair samples was published recently
832 (Salomone et al., 2012b). In this study, a UHPLC-MS/MS procedure was developed and validated
833 to determine the possible presence of five synthetic cannabinoids (i.e., consistently present in the
834 black market during the period considered: JWH-018, JWH-073, JWH 200, JWH-250, and HU-
835 210) in 179 real hair samples collected from previously proven Cannabis consumers (Figure 5). The
836 fourteen (7.82%) samples found positive to at least one synthetic cannabinoid demonstrated
837 significant diffusion among drug abusers. Concentrations of synthetic cannabinoids ranged from
838 0.50 to 730 pg/mg. Although the published screening was limited to relatively few compounds, its
839 UHPLC-MS/MS approach makes further method expansion feasible, so as to include new
840 substances that are continuously introduced into the black market.

841

842 **B. Future Trends Driven by Instrumental Developments**

843

844 After years of pioneering studies, hair analysis for toxicology applications represents nowadays
845 a reliable and well-established means of forensic investigation. The Society of Hair Testing
846 periodically meets to exchange scientific experiences, establish new protocols, and draw up
847 consensus documents to carefully guide conclusions and legal judgments on the basis of
848 experimental results from hair analysis.

849 Whereas instrumental sensitivity and overall method specificity have been for years key issues
850 of experimental studies and analytical method developments, the latest instrumentation for sample
851 treatment, chromatographic separation, and mass spectrometric detection substantially fulfills these
852 requirements, and frequently provides LOD and LOQ values largely better (i.e., lower) than the
853 expected drug concentrations in hair. Indeed, various sources of individual variability, including
854 genetic polymorphisms, metabolic disorders, diet, use of cosmetics, and environmental exposure,
855 are likely to represent more-important causes of bias and incorrect reporting than insufficient
856 sensitivity and specificity. For example, the LOD for EtG in hair will soon break the 1 pg/mg limit
857 and will open up a reflection on the possible origins of this alcohol metabolite other than ingestion
858 of alcoholic beverages. Moreover, the chance of external contamination of hair samples and extracts
859 grows with increasing method sensitivity, in turn becoming a major concern during the
860 experimental method validation.

861 Other issues, basically linked to one another, are becoming increasingly important in forensic
862 hair analysis. These issues are: multiresidue and multiclass potential, high throughput, and costs of
863 analysis. Capability of simultaneous multiclass analysis is essential in broad drug screening of
864 unknown analytes, because in most acute intoxications and post-mortem investigations the ingested
865 substance cannot be anticipated; however, this feature also contributes to dramatically decrease the
866 number of analysis for each sample, with a resultant cost lowering. Modern UHPLC-MS/MS
867 instruments with triple-quadrupole analyzers can already simultaneously detect more than one
868 hundred substances in just a few minutes, and QuEChERS-like procedures to extract hair samples
869 are rapidly spreading within toxicology laboratories. This trend will be probably pursued further in
870 the near future, up to new limits in terms of speed and number of target analytes, so as to manage
871 hundreds of samples per day. Automatically-driven high-resolution MS/MS instruments,
872 sequentially generating data-dependent CID product-ion spectra, will provide true general unknown
873 screening and will allow retrospective inquiry of previously acquired data. Improvements in mass
874 analyzers that will lead to enhanced sensitivity and mass resolution at high scanning rate will

875 possibly make these instruments preferable to triple quadrupoles for general unknown screening
876 analysis.

877 High throughput and costs of analysis are strictly connected to one another. Hair analysis is still
878 an expensive determination that prevents its generalized application to workplace and driving
879 license testing, even if it is evident that drivers and, for example, workers who carry fire-arms
880 should not be habitual drug consumers. The analysis costs are due to the labor-intensive hair-sample
881 treatment and the difficulty to test all drugs of abuse at once. Further development of direct MS
882 methods of analysis, such as DESI or MALDI-TOF, combined with simplified sample handling,
883 could possibly provide very fast and cheap drug screening of hair samples in the future. Even if the
884 sensitivity of these instrumental approaches is still low, quite high cut-off values could be accepted
885 whenever the scope of the screening is to single out only the chronic drug consumers. Also, MS-
886 imaging techniques require significant improvement, in terms of hair sample handling and
887 instrumental sensitivity, before generalized application can be found in the chronological
888 assessment of drug intake.

889 The scientific literature reports no studies on the application of isotope-ratio mass spectrometry
890 (IRMS) to hair analysis to discriminate the natural vs. exogenic origin of substances present therein.
891 The substances of interest might include doping agents, such as steroids, as well as a common rape
892 drug; i.e., gamma-hydroxybutyric acid. The reasons for this lack can be found again in the limited
893 sensitivity of the technique, the restricted sample availability, the complexity of its purification, and
894 that specificity provided by mass selection is lost in the combustion process. The chance of using
895 IRMS to detect the illegal administration of drugs over extended periods of time, as it might occur
896 for androgenic steroids, will possibly push some scientific efforts in this direction in the near future.

897 In general, the recent major improvements of MS and chromatographic instrumentation have
898 driven forensic toxicology toward previously inconceivable results. Among these studies, it is
899 impressive to note that hair analysis can nowadays provide evidence of a single administration of
900 drugs. The continuous enhancement of MS sensitivity progressively reduces the amount of hair
901 sample needed for the chemical analysis, and extends the chance to investigate newborn children, to
902 detect pre-natal drugs exposure. The scientific goal to gain toxicological information from a single
903 hair has already started, in a worthy competition with forensic genetics. Further unexpected
904 developments of MS instruments are likely to have immediate impact also in the investigations of
905 forensic toxicology, and in particular in hair analysis.

906
907
908

909 **IV. ABBREVIATIONS**

910 All abbreviations and acronyms used in this article are listed in Table 2.

REFERENCES

- Agius R, Nadulski N, Kahl HG, Schröder J, Dufaux B, Yegles M, Pragst F. 2010. Validation of a headspace solid-phase microextraction-GC-MS/MS for the determination of ethyl glucuronide in hair according to forensic guidelines. *Forensic Sci Int* 196:3–9.
- Agius R, Nadulski N, Kahl HG, Dufaux B. 2012. Ethyl glucuronide in hair – A highly effective test for the monitoring of alcohol consumption. *Forensic Sci Int* 218:10-14.
- Albermann ME, Musshoff F, Madea B. 2010. A fully validated high-performance liquid chromatography-tandem mass spectrometry method for the determination of ethyl glucuronide in hair for the proof of strict alcohol abstinence. *Anal Bioanal Chem* 396:2441–2447.
- Albermann ME, Musshoff F, Madea B. 2011. Comparison of ethyl glucuronide (EtG) and fatty acid ethyl esters (FAEEs) concentrations in hair for testing abstinence. *Anal Bioanal Chem* 400:175–181.
- Albermann ME, Musshoff F, Aengenheister L, Madea B. 2012. Investigations on the influence of different grinding procedures on measured ethyl glucuronide concentrations in hair determined with an optimized and validated LC-MS/MS method. *Anal Bioanal Chem* 403:769:776.
- Aleksa K, Walasek P, Fulga N, Kapur B, Gareri J, Koren G. 2012. Simultaneous detection of seventeen drugs of abuse and metabolites in hair using solid-phase micro extraction (SPME) with GC/MS. *Forensic Sci Int* 218:31-36.
- Álvarez I, Bermejo AM, Tabernero MJ, Fernández P, Cabarcos P, López P. 2009. Microwave-assisted extraction: a simpler and faster method for the determination of ethyl glucuronide in hair by gas chromatography-mass spectrometry. *Anal Bioanal Chem* 393:1345–1350.
- Anielski P. 2008. Hair analysis of anabolic steroids in connection with doping control – results from horse samples. *J Mass Spectrom* 43: 1001–1008.
- Barroso M, Dias M, Vieira DN, Queiroz JA, López-Rivadulla M. 2008. Development and validation of an analytical method for the simultaneous determination of cocaine and its

main metabolite, benzoylecgonine, in human hair by gas chromatography/mass spectrometry. *Rapid Commun Mass Spectrom* 22:3320-3326.

Barroso M, Dias M, Vieira DN, López-Rivadulla M, Queiroz JA. 2010a. Simultaneous quantitation of morphine, 6-acetylmorphine, codeine, 6-acetylcodeine and tramadol in hair using mixed-mode solid-phase extraction and gas chromatography-mass spectrometry. *Anal Bioanal Chem* 396:3059-3069.

Barroso M, Costa S, Dias M, Vieira DN, Queiroz JA, Lopez-Rivadulla M. 2010b. Analysis of phenylpiperazine-like stimulants in human hair as trimethylsilyl derivatives by gas chromatography-mass spectrometry. *J Chromatogr A*. 1217:6274-6280.

Bendroth P, Kronstrand R, Helander A, Greby J, Stephanson N, Krantz P. 2008. Comparison of ethyl glucuronide in hair with phosphatidylethanol in whole blood as post-mortem markers of alcohol abuse. *Forensic Sci Int* 176:76–81.

Beuck S, Möller I, Thomas A, Klose A, Schlörer N, Schänzer W, Thevis M. 2011. Structure characterisation of urinary metabolites of the cannabimimetic JWH-018 using chemically synthesised reference material for the support of LC-MS/MS-based drug testing. *Anal Bioanal Chem* 401:493-505.

Broecker S, Herre S, Wurst B, Zweigenbaum J, Pragst F. 2011. Development practical application of a CID accurate mass spectra library of more than 2500 toxic compounds for systematic toxicological analysis by LC–QTOF-MS with data dependent acquisition. *Anal Bioanal Chem* 400:101–117.

Broecker S, Herre S, Pragst F. 2012. General unknown screening in hair by liquid chromatography–hybrid quadrupole time-of-flight mass spectrometry (LC–QTOF-MS). *Forensic Sci Int* 218:68-81.

Bucelli F, Fratini A, Bavazzano P, Comodo N. 2009. Quantification of drugs of abuse and some stimulants in hair samples by liquid chromatography-electrospray ionization ion trap mass spectrometry. *J Chromatogr B* 877:3931-3936.

Cabarcos P, Hassan HM, Tabernero MJ, Scott KS. 2012. Analysis of ethyl glucuronide in hair samples by liquid chromatography-electrospray ionization-tandem mass spectrometry (LC-ESI-MS/MS). *J Appl Toxicol* doi: 10.1002/jat.1791.

- Caplan YH, Goldberger BA. 2001. Alternative specimens for workplace drug testing. *J Anal Toxicol* 25:396-399.
- Caprara DL, Klein J, Koren G. 2006. Diagnosis of fetal alcohol spectrum disorder (FASD): fatty acid ethyl esters and neonatal hair analysis. *Ann Ist Super Sanita* 42:39-45.
- Chèze M, Duffort G, Deveaux M, Pèpin G. 2005. Hair analysis by liquid chromatography-tandem mass spectrometry in toxicological investigation of drug-facilitated crimes: report of 128 cases over the period June 2003-May 2004 in metropolitan Paris. *Forensic Sci Int* 153:3-10.
- Chèze M, Deveaux M, Martin C, Lhermitte M, Pépin G. 2007. Simultaneous analysis of six amphetamines and analogues in hair, blood and urine by LC-ESI-MS/MS. Application to the determination of MDMA after low ecstasy intake. *Forensic Sci Int* 170:100-104.
- Cooper GA, Kronstrand R, Kintz P. 2012. Society of Hair Testing guidelines for drug testing in hair. *Forensic Sci Int* 218:20–24.
- Cordero R, Paterson S. 2007. Simultaneous quantification of opiates, amphetamines, cocaine and metabolites and diazepam and metabolite in a single hair sample using GC-MS. *J Chromatogr B* 850:423-431.
- Deshmukh N, Hussain I, Barker J, Petroczi A, Naughton DP. 2010. Analysis of anabolic steroids in human hair using LC-MS/MS. *Steroids* 75: 710–714.
- Deshmukh N, Barker J, Petroczi A, Naughton DP. 2012. Detection of testosterone and epitestosterone in human hair using liquid chromatography-tandem mass spectrometry. *J Pharm Biomed Anal* doi.org/10.1016/j.jpba.2012.04.011.
- Di Corcia D, D'Urso F, Gerace E, Salomone A, Vincenti M. 2012. Simultaneous determination in hair of multi-class drugs of abuse (including THC) by ultra-high performance liquid chromatography-tandem mass spectrometry. *J Chromatogr B* doi:10.1016/j.jchromb.2012.05.003.
- Domínguez-Romero JC, García-Reyes JF, Molina-Díaz A. 2011. Screening and quantitation of multi-class drugs of abuse and pharmaceuticals in hair by fast liquid chromatography electrospray time-of-flight mass spectrometry. *J Chromatogr B* 879:2034-2042.

- Dresen S, Kneisel S, Weinmann W, Zimmermann R, Auwärter V. 2011. Development and validation of a liquid chromatography-tandem mass spectrometry method for the quantitation of synthetic cannabinoids of the aminoalkylindole type and methanandamide in serum and its application to forensic samples. *J Mass Spectrom* 46:163-171.
- Emídio ES, Prata Vde M, Dórea HS. 2010. Validation of an analytical method for analysis of cannabinoids in hair by headspace solid-phase microextraction and gas chromatography-ion trap tandem mass spectrometry. *Anal Chim Acta* 670:63-71.
- EWDTS (European Workplace Drug Testing Society) Guidelines. 2010. Drug and alcohol testing in hair, collection and analysis. EWDTS, August 2010.
- Falcon M, Pichini S, Joya J, Pujadas M, Sanchez A, Vall O, García Algar O, Luna A, de la Torre R, Rotolo MC, Pellegrini M. 2012. Maternal hair testing for the assessment of fetal exposure to drug of abuse during early pregnancy: Comparison with testing in placental and fetal remains. *Forensic Sci Int* 218:92-96.
- Favretto D, Vogliardi S, Stocchero G, Nalesso A, Tuccia M, Ferrara SD. 2012. High performance liquid chromatography–high resolution mass spectrometry and micropulverized extraction for the quantification of amphetamines, cocaine, opioids, benzodiazepines, antidepressants and hallucinogens in 2.5 mg hair samples. *J Chromatography A* 1218:6583-6595.
- Gaillard Y, Vayssette F, Balland A, Pèpin G. 1999. Gas chromatographic–tandem mass spectrometric determination of anabolic steroids and their esters in hair. Application in doping control and meat quality control. *J Chromatogr B*. 735 :189–205.
- Gaillard Y, Vayssette F, Pèpin G. 2000. Compared interest between hair analysis and urinalysis in doping controls. Results for amphetamines, corticosteroids and anabolic steroids in racing cyclists. *Forensic Sci Int* 107: 361–379.
- Gambelunghe C, Somavilla M, Ferranti C, Rossi R, Aroni K, Manes N, Bacci M. 2007. Analysis of anabolic steroids in hair by GC/MS/MS. *Biomed Chromatogr* 21: 369–375.
- Gareri J, Appenzeller B, Walasek P, Koren G. 2011. Impact of hair-care products on FAEE hair concentrations in substance abuse monitoring. *Anal Bioanal Chem* 400:183–188.

- Gaulier JM, Sauvage FL, Pauthier H, Saint-Marcoux F, Marquet P, Lachâtre G. 2008. Identification of acepromazine in hair: an illustration of the difficulties encountered in investigating drug-facilitated crimes. *J Forensic Sci* 53:755-759.
- Gottardo R, Fanigliulo A, Bortolotti F, De Paoli G, Pascali JP, Tagliaro F. 2007. Broad-spectrum toxicological analysis of hair based on capillary zone electrophoresis-time-of-flight spectrometry. *J Chromatogr A* 1159:190-197.
- Gottardo R, Mikš'ík I, Aturki Z, Sorio D, Seri C, Fanali S, Tagliaro F. 2012. Analysis of drugs of forensic interest with capillary zone electrophoresis/time-of-flight mass spectrometry based on the use of non-volatile buffers. *Electrophoresis* 33:599-606.
- Grigoryev A, Savchuk S, Melnik A, Moskaleva N, Dzhurko J, Ershov M, Nosyrev A, Vedenin A, Izotov B, Zabirowa I, Rozhanets V. 2011. Chromatography-mass spectrometry studies on the metabolism of synthetic cannabinoids JWH-018 and JWH-073, psychoactive components of smoking mixtures. *J Chromatogr B* 879:1126-1136.
- Guthery B, Bassindale T, Bassindale A, Pillinger CT, Morgan GH. 2010. Qualitative drug analysis of hair extracts by comprehensive two-dimensional gas chromatography/time-of-flight mass spectrometry. *J Chromatogr A* 1217:4402-4410.
- Harun N, Anderson RA, Cormack PAG. 2010. Analysis of ketamine and norketamine in hair samples using molecularly imprinted solid-phase extraction (MISPE) and liquid chromatography-tandem mass spectrometry (LC-MS/MS). *Anal Bioanal Chem* 396:2449-2459.
- Hastedt M, Herre S, Pragst F, Rothe M, Hartwig S. 2012. Workplace alcohol testing program by combined use of ethyl glucuronide and fatty acid ethyl esters in hair. *Alcohol Alcohol* 47:127-132.
- Hegstad S, Khiabani HZ, Kristoffersen L, Kunøe N, Lobmaier PP, Christophersen AS. 2008. Drug screening of hair by liquid chromatography-tandem mass spectrometry. *J Anal Toxicol* 32:364-372.
- Huang DK, Liu C, Huang MK, Chien CS. 2009. Simultaneous determination of morphine, codeine, 6-acetylmorphine, cocaine and benzoyllecgonine in hair by liquid

chromatography/electrospray ionization tandem mass spectrometry. *Rapid Commun Mass Spectrom* 23:957-962.

Irving RC, Dickson SJ. 2007. The detection of sedatives in hair and nail samples using tandem LC-MS-MS. *Forensic Sci Int* 166:58-67.

Jackson AU, Garcia-Reyes JF, Harper JD, Wiley JS, Molina-Díaz A, Ouyang Z, Cooks RG. 2010. Analysis of drugs of abuse in biofluids by low temperature plasma (LTP) ionization mass spectrometry. *Analyst* 135:927-933.

Johansen SS, Dahl-Sørensen R. 2012. A drug rape case involving triazolam detected in hair and urine. *Int J Legal Med* doi:10.1007/s00414-011-0654-6.

Kerekes I, Yegles M, Grimm U, Wennig R. 2009. Ethyl glucuronide determination: head hair versus non-head hair. *Alcohol Alcohol* 44:62–66.

Kharbouche H, Sporkert F, Troxler S, Augsburger M, Mangin P, Staub C. 2009. Development and validation of a gas chromatography-negative chemical ionization tandem mass spectrometry method for the determination of ethyl glucuronide in hair and its application to forensic toxicology. *J Chromatogr B* 877:2337–2343.

Kharbouche H, Steiner N, Morelato M, Staub C, Boutrel B, Mangin P, Sporkert F, Augsburger M. 2010. Influence of ethanol dose and pigmentation on the incorporation of ethyl glucuronide into rat hair. *Alcohol* 44:507-514.

Kim J, Lee S, In S, Choi H, Chung H. 2011. Validation of a simultaneous analytical method for the detection of 27 benzodiazepines and metabolites and zolpidem in hair using LC-MS/MS and its application to human and rat hair. *J Chromatogr B* 879:878-886.

King LA, Kicman AT. 2011. A brief history of ‘new psychoactive substances’. *Drug Test Anal* 3:401–403.

Kintz P, Cirimele V, Ludes B. 2000. Pharmacological criteria that can affect the detection of doping agents in hair. *Forensic Sci Int* 107: 325–334.

Kintz P. 2003. Testing for anabolic steroids in hair: a review. *Leg Med (Tokyo)* 5:S29–S33.

- Kintz P, Villain M, Dumestre-Toulet V, Ludes B. 2005. Drug-facilitated sexual assault and analytical toxicology: the role of LC-MS/MS A case involving zolpidem. *J Clin Forensic Med* 12:36-41.
- Kintz P, Villain M, Cirimele V. 2006. Hair analysis for drug detection. *Ther Drug Monit* 28:442-446.
- Kintz P. 2007a. Analytical and practical aspects of drug testing in hair. Boca Raton: CRC Press. Taylor & Francis 382 p.
- Kintz P. 2007b. Bioanalytical procedures for detection of chemical agents in hair in the case of drug-facilitated crimes. *Anal Bioanal Chem* 388:1467-1474.
- Kintz P, Villain M, Vallet E, Etter M, Salquebre G, Cirimele V. 2008. Ethyl glucuronide: unusual distribution between head hair and pubic hair. *Forensic Sci Int* 176:87–90.
- Kintz P, Evans J, Villain M, Chatterton C, Cirimele V. 2009. Hair analysis to demonstrate administration of sildenafil to a woman in a case of drug-facilitated sexual assault. *J Anal Toxicol* 33:553-556.
- Kintz P. 2012a. Value of the concept of minimal detectable dosage in human hair. *Forensic Sci Int* 218:28-30.
- Kintz P. 2012b. Consensus of the Society of Hair Testing on hair testing for chronic excessive alcohol consumption 2011. *Forensic Sci Int* 218:2.
- Klis M, Rojek S, Kulikowska J, Bożek E, Ścisłowski M. 2007. Usefulness of multi-parameter opiates-amphetamines-cocainics analysis in hair of drug users for the evaluation of an abuse profile by means of LC-APCI-MS-MS. *J Chromatogr B* 854:299-307.
- Kraemer T, Rust KY, Meyer MR, Wissenbach DK, Bregel D, Hopf M, Maurer HH, Wilske J. 2009. Studies on the metabolism of JWH-018 and of a homologue of CP 47,497, pharmacologically active ingredients of different misused incense (“spice”) using GC–MS and LCMSn techniques. *Ann Toxicol Anal* 21:S1-S21.
- Kronstrand R, Brinkhagen L, Nyström FH. 2012. Ethyl glucuronide in human hair after daily consumption of 16 or 32 g of ethanol for 3 months. *Forensic Sci Int* 215:51–55.

- Kulaga V, Caprara D, Iqbal Q, Kapur B, Klein J, Reynolds J, Brien J, Koren G. 2006. Fatty acid ethyl esters (FAEE); comparative accumulation in human and guinea pig hair as a biomarker for prenatal alcohol exposure. *Alcohol Alcohol* 41:534–539.
- Kulaga V, Velazquez-Armenta Y, Aleksa K, Vergee Z, Koren G. 2009. The effect of hair pigment on the incorporation of fatty acid ethyl esters (FAEE). *Alcohol Alcohol* 44:287–292.
- Laloup M, Ramirez Fernandez MM, Wood M, Maes V, De Boeck G, Vanbeckevoort, Samyn N. 2007. Detection of diazepam in urine, hair and preserved oral fluid samples with LC-MS-MS after single and repeated administration of Myolastan and Valium. *Anal Bioanal Chem* 388:1545-1556.
- Lamoureux F, Gaulier JM, Sauvage FL, Mercerolle M, Vallejo C, Lachâtre G. 2009. Determination of ethyl-glucuronide in hair for heavy drinking detection using liquid chromatography-tandem mass spectrometry following solid-phase extraction. *Anal Bioanal Chem* 394:1895–1901.
- LeBeau MA, Mozayani A. 2001. Collection of evidence from DFSA. In: LeBeau MA, Mozayani A, Editor. *Drug-Facilitated Sexual Assault: a forensic handbook*. San Diego: Academic Press. p 197-209.
- LeBeau MA, Montgomery MA. 2009. Considerations on the utility of hair analysis for cocaine. *J Anal Toxicol* 33: 343-344.
- LeBeau MA, Montgomery MA. 2010. Challenges of Drug-Facilitated Sexual Assault. *Forensic Sci Rev* 22:1-6.
- Lee S, Han E, Park Y, Choi H, Chung H. 2009. Distribution of methamphetamine and amphetamine in drug abusers' head hair. *Forensic Sci Int* 190:16-18.
- Lee S, Kim J, In S, Choi H, Oh SM, Jang CG, Chung KH. 2012. Development of a simultaneous analytical method for selected anorectics, methamphetamine, MDMA, and their metabolites in hair using LC-MS/MS to prove anorectics abuse. *Anal Bioanal Chem* 403:1385-1394.
- Lendoiro E, Quintela O, de Castro A., Cruz A, Lopez-Rivadulla M, Concheiro M. 2012. Target screening and confirmation of 35 licit and illicit drugs and metabolites by LC-MSMS. *Forensic Sci Int* 217:207-215.

- Lin YH, Lee MR, Lee RJ, Ko WK, Wu SM. 2007. Hair analysis for metamphetamine, ketamine, morphine and codeine by cation-selective exhaustive injection and sweeping electrokinetic chromatography. *J. Chromatogr. A* 1145:234-240.
- Liu HC, Liu RH, Lin DL, Ho HO. 2010. Rapid screening and confirmation of drugs and toxic compounds in biological specimens using liquid chromatography/ion trap tandem mass spectrometry and automated library search. *Rapid Commun Mass Spectr* 24:75-84.
- Marques P, Tippetts S, Allen J, Javors M, Alling C, Yegles M, Pragst F, Wurst F. 2010. Estimating driver risk using alcohol biomarkers, interlock blood alcohol concentration tests and psychometric assessments: initial descriptives. *Addiction* 105:226-239.
- Martin M, Muller JF, Turner K, Duez M, Cirimele V. 2012. Evidence of mephedrone chronic abuse through hair analysis using GC/MS. *Forensic Sci Int* 218:44-48.
- Matuszewski BK, Constanzer ML, Chavez-Eng CM. 2003. Strategies for the assessment of matrix effect in quantitative bioanalytical methods based on HPLC-MS/MS. *Anal Chem* 75:3019-3030.
- Merola G, Gentili S, Tagliaro F, Macchia T. 2010. Determination of different recreational drugs in hair by HS-SPME and GC/MS. *Anal Bioanal Chem* 397:2987:2995.
- Miki A, Katagi M, Kamata T, Zaitse K, Tatsuno M, Nakanishi T, Tsuchihashi H, Takubo T, Suzuki K. 2011. MALDI-TOF and MALDI-FTICR imaging mass spectrometry of methamphetamine incorporated into hair. *J Mass Spectrom* 46:411-416.
- Miller EI, Wylie FM, Oliver JS. 2008. Simultaneous detection and quantification of amphetamines, diazepam and its metabolites, cocaine and its metabolites, and opiates in hair by LC-ESI-MS-MS using a single extraction method. *J Anal Toxicol* 32:457-469.
- Miyaguchi H, Kakuta M, Iwata YT, Matsuda H, Tazawa H, Kimura H, Inoue H. 2007. Development of a micropulverized extraction method for rapid toxicological analysis of methamphetamine in hair. *J Chromatogr A* 1163:43-48.
- Miyaguchi H, Iwata YT, Kanamori T, Tsujikawa K, Kuwayama K, Inoue H. 2009. Rapid identification and quantification of methamphetamine and amphetamine in hair by gas

- chromatography/mass spectrometry coupled with micropulverized extraction, aqueous acetylation and microextraction by packed sorbent. *J Chromatogr A* 1216:4063-4070.
- Moore C, Coulter C, Crompton K. 2007. Determination of cocaine, benzoylecgonine, cocaethylene and norcocaine in human hair using solid-phase extraction and liquid chromatography with tandem mass spectrometric detection. *J Chromatogr B* 859:208-212.
- Moran CL, Le VH, Chimalakonda KC, Smedley AL, Lackey FD, Owen SN, Kennedy PD, Endres GW, Ciske FL, Kramer JB, Kornilov AM, Bratton LD, Dobrowolski PJ, Wessinger WD, Fantegrossi WE, Prather PL, James LP, Radominska-Pandya A, Moran JH. 2011. Quantitative Measurement of JWH-018 and JWH-073 Metabolites Excreted in Human Urine. *Anal Chem* 83:4228-4236.
- Morini L, Politi L, Poletti A. 2009. Ethyl glucuronide in hair. A sensitive and specific marker of chronic heavy drinking. *Addiction* 104:915-920.
- Morini L, Zucchella A, Poletti P, Politi L, Groppi A. 2010. Effect of bleaching on ethyl glucuronide in hair: an in vitro experiment. *Forensic Sci Int* 198:23-27.
- Nadulski T, Pragst F. 2007. Simple and sensitive determination of Delta(9)-tetrahydrocannabinol, cannabidiol and cannabinol in hair by combined silylation, headspace solid-phase microextraction and gas chromatography-mass spectrometry. *J Chromatogr B* 846:78-85.
- Negrusz A, Moore CM, Kern JL, Janicak PG, Strong MJ, Levy NA. 2000. Quantitation of clonazepam and its major metabolite 7-aminoclonazepam in hair. *J Anal Toxicol* 24:614-620.
- Negrusz A, Moore CM, Hinkel KB, Stockham TL, Verma M, Strong MJ, Janicak PG. 2001. Deposition of 7-aminoflunitrazepam and flunitrazepam in hair after a single dose of Rohypnol. *J Forensic Sci* 46:1143-1151.
- Negrusz A, Gaensslen RE. 2003. Analytical developments in toxicological investigation of drug-facilitated sexual assault. *Anal Bioanal Chem* 376:1192-1197.
- Nielen MWF, Nijroldera AWJM, Hooijerinka H, Stolkera AAM. 2011. Feasibility of desorption electrospray ionization mass spectrometry for rapid screening of anabolic steroid esters in hair. *Anal Chim Acta* 700:63-69.

- Nielsen MK, Johansen SS, Dalsgaard PW, Linnet K. 2010. Simultaneous screening and quantification of 52 common pharmaceuticals and drugs of abuse in hair using UPLC-TOF-MS. *Forensic Sci Int* 196:85-92.
- Niessen WMA. 2011. Fragmentation of toxicologically relevant drugs in positive-ion liquid chromatography–tandem mass spectrometry. *Mass Spectrom Rev* 30:626-663.
- Parkin MC, Brailsford AD. 2009. Retrospective drug detection in cases of drug-facilitated sexual assault: challenges and perspectives for the forensic toxicologist. *Bioanalysis* 1:1001-1013.
- Paul R, Kingston R, Tsanaclis L, Berry A, Guwy A. 2008. Do drug users use less alcohol than non-drug users? A comparison of ethyl glucuronide concentrations in hair between the two groups in medico-legal cases. *Forensic Sci Int* 176:82–86.
- Paul R, Tsanaclis L, Kingston R, Berry A, Guwy A. 2011. Simultaneous determination of GHB and EtG in hair using GCMS/MS. *Drug Test Anal* 3:201-205.
- Pelander A, Ristimaa J, Rasanen I, Vuori E, Ojanperä I. 2008. Screening for basic drugs in hair of drug addicts by liquid chromatography/time-of-flight mass spectrometry. *Ther Drug Monit* 30:717-724.
- Peters FT, Remane D. 2012. Aspects of matrix effects in applications of liquid chromatography-mass spectrometry to forensic and clinical toxicology-a review. *Anal Bioanal Chem* doi: 10.1007/s00216-012-6035-2.
- Pirro V, Valente V, Oliveri P, De Bernardis A, Salomone A, Vincenti M. 2011a. Chemometric evaluation of nine alcohol biomarkers in a large population of clinically-classified subjects: pre-eminence of ethyl glucuronide concentration in hair for confirmatory classification. *Anal Bioanal Chem* 401:2153-2164.
- Pirro V, Di Corcia D, Pellegrino S, Vincenti M, Sciutteri B, Salomone A. 2011b. A study of distribution of ethyl glucuronide in different keratin matrices. *Forensic Sci Int* 210:271–277.
- Polettini A, Cone EJ, Gorelick DA, Huestis MA. 2012. Incorporation of methamphetamine and amphetamine in human hair following controlled oral methamphetamine administration. *Anal Chim Acta* 726:35-43.

- Politi L, Morini L, Leone F, Polettini A. 2006. Ethyl glucuronide in hair: is it a reliable marker of chronic high levels of alcohol consumption?. *Addiction* 101:1408–1412.
- Politi L, Morini L, Mari F, Groppi A, Bertol E. 2008. Ethyl glucuronide and ethyl sulfate in autopsy samples 27 years after death. *Int J Leg Med* 122:507-509.
- Politi L, Mari F, Furlanetto S, Del Bravo E, Bertol E. 2011. Determination of fatty acid ethyl esters in hair by GC-MS and application in a population of cocaine users. *J Pharm Biomed Anal* 54:1192–1195.
- Porta T, Grivet C, Kraemer T, Varesio E, Hopfgartner G. 2011. Single hair cocaine consumption monitoring by mass spectrometry imaging. *Anal Chem* 83:4266-4272.
- Pozo OJ, Deventer K, Van Eenoo P, Rubens R, Delbeke FT. 2009. Quantification of testosterone undecanoate in human hair by liquid chromatography-tandem mass spectrometry. *Biomed Chromatogr* 23: 873–880.
- Pragst F, Auwaerter V, Sporkert F, Spiegel K. 2001. Analysis of fatty acid ethyl esters in hair as possible markers of chronically elevated alcohol consumption by headspace solid-phase microextraction (HS-SPME) and gas chromatography-mass spectrometry (GC-MS). *Forensic Sci Int* 121:76-88.
- Pragst F, Nadulski T. 2005. Limite de détection du THC dans les cheveux dans le contexte de l'aptitude à la conduite. *Ann Toxicol Anal* 17:237-240.
- Pragst F, Balikova MA. 2006. State of the art in hair analysis for detection of drug and alcohol abuse. *Clin Chim Acta* 370:17-49.
- Pragst F, Yegles M. 2008. Determination of fatty acid ethyl esters (FAEE) and ethyl glucuronide (EtG) in hair: a promising way for retrospective detection of alcohol abuse during pregnancy?. *Ther Drug Monit* 30:255-263.
- Pragst F, Rothe M, Moench B, Hastedt M, Herre S, Simmert D. 2010. Combined use of fatty acid ethyl esters and ethyl glucuronide in hair for diagnosis of alcohol abuse: interpretation and advantages. *Forensic Sci Int* 196:101–110.

- Quintela O, Lendoiro E, Cruz A, de Castro A, Quevedo A, Jurado C, López-Rivadulla M. 2010. Hydrophilic interaction liquid chromatography-tandem mass spectrometry (HILIC-MS/MS) determination of cocaine and its metabolites benzoylecgonine, ecgonine methyl ester, and cocaethylene in hair samples. *Anal Bioanal Chem* 396:1703-1712.
- Rodrigues de Oliveira CD, Yonamine M, and de Moraes Moreau RL 2007. Headspace solid-phase microextraction of cannabinoids in human head hair samples. *J Sep Sci* 30:128-134.
- Salomone A, Gerace E, Di Corcia D, Martra G, Petrarulo M, Vincenti M. 2012a. Hair analysis of drugs involved in drug-facilitated sexual assault and detection of zolpidem in a suspected case. *Int J Legal Med* 126:451-459.
- Salomone A, Gerace E, D'Urso F, Di Corcia D, Vincenti M. 2012b. Simultaneous analysis of several synthetic cannabinoids, THC, CBD, CBN in hair by ultra-high performance liquid chromatography tandem mass spectrometry. Method validation and application to real samples. *J Mass Spectrom* 47:604–610.
- Scott KS. 2009. The use of hair as a toxicological tool in DFC casework. *Sci Justice* 49:250-253.
- Shen M, Xiang P, Shen B, Bu J, Wang M. 2009. Physiological concentrations of anabolic steroids in human hair. *Forensic Sci Int* 184: 32–36.
- Shi Y, Shen B, Xiang P, Yan Y, Shen M. 2010. Determination of ethyl glucuronide in hair samples of Chinese people by protein precipitation (PPT) and large volume injection-gas chromatography-tandem mass spectrometry (LVI-GC/MS/MS). *J Chromatogr B* 878:3161–3166.
- Sobolevsky T, Prasolov I, Rodchenkov G. 2010. Detection of JWH-018 metabolites in smoking mixture post-administration urine. *Forensic Sci Int* 200:141-147.
- Sporkert F, Kharbouche H, Augsburg MP, Klemm C, Baumgartner MR. 2012. Positive EtG findings in hair as a result of a cosmetic treatment. *Forensic Sci Int* 218:97–100.
- Stout PA, Simons KD, Kerrigan S. 2010. Quantitative analysis of gamma-hydroxybutyrate at endogenous concentrations in hair using liquid chromatography tandem mass spectrometry. *J Forensic Sci* 55:531-537.

- Süsse S, Selavka CM, Mieczkowski T, Pragst F. 2010. Fatty acid ethyl ester concentrations in hair and self-reported alcohol consumption in 644 cases from different origin. *Forensic Sci Int* 196:111-117.
- Süsse S, Pragst F, Mieczkowski T, Selavka CM, Elian A, Sachs H, Hastedt M, Rothe M, Campbell J. 2012. Practical experiences in application of hair fatty acid ethyl esters and ethyl glucuronide for detection of chronic alcohol abuse in forensic cases. *Forensic Sci Int* 218:82-91.
- Tagliaro F, Bortolotti F, Viel G, Ferrara SD. 2011. Caveats against an improper use of hair testing to support the diagnosis of chronic excessive alcohol consumption, following the "Consensus" of the Society of Hair Testing 2009 [*Forensic Science International* 196 (2010) 2]. *Forensic Sci Int* 207:e69-e70.
- Takáts Z, Wiseman JM, Gologan B, Cooks RG. 2004. Mass spectrometry sampling under ambient conditions with desorption electrospray ionization. *Science* 306:471-473.
- Tarcomnicu I, van Nuijs AL, Aerts K, De Doncker M, Covaci A, Neels H. 2010. Ethyl glucuronide determination in meconium and hair by hydrophilic interaction liquid chromatography-tandem mass spectrometry. *Forensic Sci Int* 196:121-127.
- Teske J, Weller JP, Fieguth A, Rothämel T, Schulz Y, Tröger HD. 2010. Sensitive and rapid quantification of the cannabinoid receptor agonist naphthalen-1-yl-(1-pentylindol-3-yl) methanone (JWH-018) in human serum by liquid chromatography-tandem mass spectrometry. *J Chromatogr B* 878:2959-2963.
- Thibert V, Legeay P, Chapuis-Hugon F, Pichon V. 2012. Synthesis and characterization of molecularly imprinted polymers for the selective extraction of cocaine and its metabolite benzoylecgonine from hair extract before LC-MS analysis. *Talanta* 88:412-419.
- Thieme D, Grosse J, Sachs H, Mueller RK. 2000. Analytical strategy for detecting doping agents in hair. *Forensic Sci Int* 107: 335-345.
- Thieme D, Sachs H. 2007. Examination of a long-term clozapine administration by high resolution segmental hair analysis. *Forensic Sci Int* 166:110-114.

- Torrance H, Cooper G. 2010. The detection of mephedrone (4-methylmethcathinone) in 4 fatalities in Scotland. *Forensic Sci Int* 202:e62-e63.
- UNODC (United Nations Office on Drug and Crime). 2011. Guidelines for the Forensic analysis of drugs facilitating sexual assault and other criminal acts. UNODC, December 2011.
- Villain M, Chèze M, Tracqui A, Ludes B, Kintz P. 2004. Windows of detection of zolpidem in urine and hair: application to two drug facilitated sexual assaults. *Forensic Sci Int* 143:157-161.
- Villain M, Concheiro M, Cirimele V, Kintz P. 2005. Screening method for benzodiazepines and hypnotics in hair at pg/mg level by liquid chromatography-mass spectrometry/mass spectrometry. *J Chromatogr B* 825:72-78.
- Vogliardi S, Favretto D, Frison G, Ferrara SD, Seraglia R, Traldi P. 2009. A fast screening MALDI method for the detection of cocaine and its metabolites in hair. *J Mass Spectrom* 44:18–24.
- Vogliardi S, Favretto D, Frison G, Maietti S, Viel G, Seraglia R, Traldi P, Ferrara SD. 2010. Validation of a fast screening method for the detection of cocaine in hair by MALDI-MS. *Anal Bioanal Chem* 396:2435-2440.
- Vogliardi S, Favretto D, Tucci M, Stocchero G, Ferrara SD. 2011. Simultaneous LC-HRMS determination of 28 benzodiazepines and metabolites in hair. *Anal Bioanal Chem* 400:51-67.
- Wada M, Ikeda R, Kuroda N, Nakashima K. 2010. Analytical methods for abused drugs in hair and their applications. *Anal Bioanal Chem* 397:1039-1067.
- WADA (World Anti-Doping Agency). 2012. International standard for laboratories, Version 7.0. WADA, January 2012.
- Wintermeyer A, Möller I, Thevis M, Jübner M, Beike J, Rothschild MA, Bender K. 2010. In vitro phase I metabolism of the synthetic cannabimimetic JWH-018. *Anal Bioanal Chem* 398:2141-2153.
- Wu YH, Lin KL, Chen SC, Chang YZ. 2008a. Simultaneous quantitative determination of amphetamines, ketamine, opiates and metabolites in human hair by gas chromatography/mass spectrometry. *Rapid Commun Mass Spectrom* 22:887-897.

- Wu YH, Lin KL, Chen SC, Chang YZ. 2008b. Integration of GC/EI-MS and GC/NCI-MS for simultaneous quantitative determination of opiates, amphetamines, MDMA, ketamine, and metabolites in human hair. *J Chromatogr B* 870:192-202.
- Xiang P, Sun Q, Shen B, Chen P, Liu W, Shen M. 2011. Segmental hair analysis using liquid chromatography–tandem mass spectrometry after a single dose of benzodiazepines. *Forensic Sci Int* 204:19-26.
- Zimmermann CM, Jackson GP. 2010. Gas chromatography tandem mass spectrometry for biomarkers of alcohol abuse in human hair. *Ther Drug Monit* 32:216-223.

TABLE 1. Comparison of MS conditions for EtG determination in hair samples

Authors, year	Analytical Technique	Mass Analyzer	<i>m/z</i> Ions (acquisition mode)	Derivatization
Álvarez et al., 2009	GC-EI-MS	single quadrupole	261; 160; 405 (SIM)	BSTFA
Paul et al., 2008 Paul et al., 2011	GC-EI-MS/MS	triple quadrupole	261 → 143 (SRM)	BSTFA
Shi et al., 2010	GC-EI-MS/MS	triple quadrupole	261 → 143 261 → 73 (SRM)	BSTFA
Kharbouche et al., 2009 Kharbouche et al., 2010 Sporkert et al., 2012	GC-NCI-MS/MS	triple quadrupole	347 → 163 347 → 119 (SRM)	PFPA
Agius et al., 2010	GC-NCI-MS/MS	triple quadrupole	596 → 427 596 → 288 (SRM)	HFBA
Kerekes et al., 2009	GC-NCI-MS	single quadrupole	496; 349 (SIM)	PFPA
Morini et al., 2009 Morini et al., 2010 Politi et al., 2008	LC-ESI-MS/MS	triple quadrupole	221 → 221 221 → 85 221 → 75 (SRM)	-
Pirro et al., 2011a Pirro et al., 2011b Kronstrand et al., 2012 Pragst et al., 2010 Hastedt et al., 2012 Lamoreux et al., 2009 Tarcumnicu et al., 2010 Kintz et al., 2008	LC-ESI-MS/MS	triple quadrupole	221 → 85 221 → 75 (SRM)	-
Süsse et al., 2010 Süsse et al., 2012	LC-ESI-MS/MS	triple quadrupole	221 → 75 221 → 57 221 → 55 (SRM)	-
Cabarcos et al., 2012	LC-ESI-MS/MS	ion trap	221 → 203 221 → 85 221 → 75 (SRM)	-
Albermann et al., 2010 Albermann et al., 2012	LC-ESI-MS/MS	triple quadrupole	221 → 113 221 → 85 221 → 75 (SRM)	-

TABLE 2. Abbreviations and acronyms used in this article.

6-MAM	6-monoacetylmorphine
AASs	Anabolic Androgenic Steroids
AP	Amphetamine
APCI	Atmospheric Pressure Chemical Ionization
BSTFA	N,O-bis[trimethylsilyl]trifluoroacetamide
BZE	Benzoylecgonine
CBN	Cannabinol
CDB	Cannabidiol
CDT	Carbohydrate-Deficient Transferrin
CI	Chemical Ionization
CID	Collision-Induced Dissociation
CNS	Central Nervous System
CZE	Capillary Zone Electrophoresis
DESI	Desorption Ionization Mass Spectrometry
DFC	Drug Facilitated Crime
DFSA	Drug Facilitated Sexual Assault
DHEA	Dehydroepiandrosterone
E12	Ethyl laureate
E14	Ethyl myristate
E16	Ethyl palmitate
E16:1	Ethyl palmitoleate
E18	Ethyl stearate
E18:1	Ethyl oleate
EI	Electron Impact
ELISA	Enzyme-Linked ImmunoSorbent Assay
EME	Ecgonine Methyl Ester
ESI	Electrospray Ionization
EtG	Ethyl Glucuronide
EtS	Ethyl Sulfate
FAEE	Fatty Acid Ethyl Esters
FTICR	Fourier Transform Ion Cyclotron Resonance
GC	Gas Chromatography
HFBA	Heptafluorobutyric Anhydride
HILIC	Hydrophilic Interaction Liquid Chromatography
HPLC	High Pressure Liquid Chromatography
HRMS	High Resolution Mass Spectrometry
HS-SPME	Head Space Solid-phase MicroExtraction
IMS	Imaging Mass Spectrometry
IRMS	Isotope-Ratio Mass Spectrometry
LC	Liquid Chromatography
LLE	Liquid-Liquid Extraction
LOD	Limit of Detection
LOQ	Limit of Quantification

TABLE 2. (Continued)

LVI	Large Volume Injection
LTP	Low Temperature Plasma
MALDI	Matrix-Assisted Laser Desorption Ionization
MBDB	N-methyl-1-(3,4-methylenedioxyphenyl)-2-butanamine
MBTFA	N-methyl-bis trifluoroacetamide
MDA	3,4-methylenedioxyamphetamine
MDEA	3,4-methylenedioxyethamphetamine
MDMA	3,4-methylenedioxymethamphetamine
MiAMi	Micropulverized Extraction Aqueous Acetylation Microextraction
MIP	Molecularly Imprinted Polimers
MISPE	Molecularly Imprinted Solid-Phase MicroExtraction
MS	Mass Spectrometry
MSTFA	N-methyl-N-trimethylsilyltrifluoroacetamide
MTBSTFA	N-methyl-N-(tert-butyldimethyl)trifluoroacetamide
NCI	Negative Chemical Ionization
PEth	Phosphatidylethanol
PTV	Programmable Temperature Vaporization
QqQ	Triple quadrupole mass spectrometer
QqQLIT	Hybrid triple quadrupole linear ion trap mass spectrometer
QuEChERS	Quick Easy Cheap Effective Rugged Safe
SIM	Selected Ion Monitoring
SOHT	Society of Hair Testing
SOHT	Society of Hair Testing
SPE	Solid-phase Extraction
SRM	Selected Reaction Monitoring
THC	Δ^9 -tetrahydrocannabinol
TMCS	Trimethylchlorosilane
TOF	Time of Flight
UHPLC	Ultra High Performance Liquid Chromatography
WADA	World Anti-Doping Agency

FIGURE CAPTIONS

FIGURE 1. MALDI-MS analytical workflow for single hair analysis. The whole process from sample preparation to SRM relative quantitation and MS/MS and MS³ confirmatory analyses takes approximately 2 h, including the analysis of multiple samples (Porta et al., 2011; reproduced with permission, American Chemical Society, copyright 2011).

FIGURE 2. GC×GC contour plot (TIC) obtained for a hair sample. (1) Cotinine; (2) meconin; (3) mesocain; (4) tropacaine; 5.EDDP; (6) phenazocine; (7) hydroxycotinine-TBDMS; (8) methadone; (9) quaalude; (10) norcocaine; (11) cocaine; (12) cocaethylene; (13) codeine; (14) diazepam; (15) benzoylecgonine-TBDMS; (16) acetylcodeine; (17) cinnamoylcocaine; (18) diacetylmorphine; (19) codeine-TBDMS; (20) morphine-TBDMS; (21) 6-MAM-TBDMS; (22) Papaverine; (23) morphine-di-TBDMS (Guthery et al., 2010; reproduced with permission, Elsevier, copyright 2010).

FIGURE 3. Concentration of head hair EtG (pg/mg) versus %CDT for 175 patients clinically classified as non-drinkers (A, green dots, N=65), social drinkers (S, yellow dots, N=51), and active heavy drinkers (H, red dots, N=59). The data are reported on a logarithmic scale. Horizontal and vertical lines divide the plane into four quadrants (I-IV) that identify the number of true and false positive (or negative) results in each category: For heavy drinkers (red dots): I. True-positive results; II. false-negative results for CDT only; III. false-negative results; IV. false-negative results for head hair EtG only (Pirro et al., 2011; redrawn with permission, Springer, copyright 2011).

FIGURE 4. Proposed fragmentation pattern for testosterone undecanoate (Pozo et al., 2009; reproduced with permission, John Wiley and Sons, copyright 2009).

FIGURE 5. SRM chromatogram of a blank hair sample fortified with the target compounds: (1) JWH-200, (2) JWH-250, (3) JWH-073, (4) CBD, (5) JWH-018, (6) CBN, (7) HU-210, (8) THC, (ISTD) JWH-018-*d*9 (Salomone et al., 2012; reproduced with permission, John Wiley and Sons, copyright 2012).