



AperTO - Archivio Istituzionale Open Access dell'Università di Torino

Comment on: "Safety of Small Circular Staplers in Double Stapling Technique Anastomosis for Sigmoid Colon and Rectal Surgery"

This is the author's manuscript

Original Citation:

Availability:

This version is available http://hdl.handle.net/2318/1835450 since 2022-10-17T10:32:38Z

Terms of use:

Open Access

Anyone can freely access the full text of works made available as "Open Access". Works made available under a Creative Commons license can be used according to the terms and conditions of said license. Use of all other works requires consent of the right holder (author or publisher) if not exempted from copyright protection by the applicable law.

(Article begins on next page)

Diseases of the Colon & Rectum Publish Ahead of Print

DOI: 10.1097/DCR.00000000002402

LETTER

Comment on: "Safety of Small Circular Staplers in Double Stapling Technique Anastomosis

for Sigmoid Colon and Rectal Surgery"

Andrea Resegotti, M.D.

Marco Allaix, M.D.

Massimiliano Mistrangelo, M.D.

Mario Morino, M.D., A.S.A. (Hon.), A.F.C. (Hon.)

We read with great interest the paper by Nagaoka et al,¹ showing the safety of using 25-mm circular staplers in the double stapling technique anastomosis.

In 2016, we² published a series of 37 patients who underwent restorative proctocolectomy with double-stapled ileal-pouch-anal anastomosis using 25-mm circular staplers. Functional results were good. Mean number of bowel movements was 4.5 during the day and 0.9 at night, 68% of patients were fully continent, and 32% had occasional soiling. No one reported incontinence: we ascribed those results to reduced sphincter injury and functional damage during transanal introduction of a smaller stapler.

Additional advantages are a reduced risk of entrapment of the vagina during stapler closure and, in the double stapling technique, a reduced risk of disruption of the linear staple line when the circular stapler is introduced in a very short anal stump.

That a smaller stapler could increase the risk of anastomotic stenosis is far from demonstrated; even if so, gentle digital dilatation in the office is usually effective, and it is straightforward in a very low anastomosis.

So, we agree with Nagaoka et al about the advantages of small-diameter circular staplers; in our opinion, the 25-mm circular stapler should be the stapler of choice for ileal-pouch anal anastomosis.



References

- Nagaoka T, Yamaguchi T, Nagasaki T, et al. Safety of small circular staplers in double stapling technique anastomosis for sigmoid colon and rectal cancer. *Dis Colon Rectum*. 2021;64:937– 945.
- 2. Resegotti A, Silvestri S, Astegiano M, et al. "Small is beautiful" A series of ileo-anal anastomoses performed with the 25 mm circular stapler. *Ann Ital Chir*. 2016;87:186–191.